Test Software P/N 4102483 Tested Part Number LA1054096

Test Software Name Trilogy FSA (PM TEST)

Software Revision 24.0.0.0 Prepared By P. Pascal

PHILIPS RESPIRONICS

# Test Report

| Step | Description | Units | Limits | Results | Pass | Fail |
|---|---|---|---|---|---|---|
| 71 | 0030.1030 Contr.Press.Ver.(36153): P1 at 46,11 (SP 46, 1, | cmH2O | 45,11 to 47,11 | 46,00 | ✓ | |
| 72 | 0030.1040 Monit.Press.Ver.(35786): P2 at 46,11 (SP 46, 1, | cmH2O | 45,11 to 47,11 | 46,02 | ✓ | |
| 73 | 0030.0164 Pprox Press. Ver.(36276): P2 at 46,11 (SP 46, 1 | cmH2O | 45,11 to 47,11 | 45,84 | ✓ | |
| 74 | 0030.1050 Contr.Press.Ver.(38571): P1 at 50,15 (SP 50, 1, | cmH2O | 49,15 to 51,15 | 49,99 | ✓ | |
| 75 | 0030.1060 Monit.Press.Ver.(38617): P2 at 50,15 (SP 50, 1, | cmH2O | 49,15 to 51,15 | 50,06 | ✓ | |
| 76 | 0030.0165 Pprox Press. Ver.(38643): P2 at 50,15 (SP 50, 1 | cmH2O | 49,15 to 51,15 | 49,90 | ✓ | |
| 77 | 0030.1070 Contr.Press.Ver.(46797): P1 at 63,95 (SP 65, 1, | cmH2O | 62,95 to 64,95 | 63,62 | ✓ | |
| 78 | 0030.1080 Monit.Press.Ver.(48203): P2 at 63,95 (SP 65, 1, | cmH2O | 62,95 to 64,95 | 63,82 | ✓ | |
| 79 | 0030.0166 Pprox Press. Ver.(46847): P2 at 63,95 (SP 65, 1 | cmH2O | 62,95 to 64,95 | 63,49 | ✓ | |
| 80 | 0030.1150 Atmospheric Pressure Verify: Patm at 28,05,28,0 | inHg | 27,45 to 28,65 | 28,13 | ✓ | |
| 81 | 0030.1250 Airstream Temperature (Tair) at 25,00 [25m 28s] | C deg | 20,00 to 40,00 | 29,75 | ✓ | |

Test Started On 01/17/24 01:38:30

Serial Number TV114090503 Elapsed Test Time 25m 28s

Status **FAIL**



Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Production mode.
File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20