Test Software P/N 4102483 **Tested Part Number** 1040000

Test Software Name Trilogy FSA (REPAIR TEST)

Software Revision 24.0.0.0 **Prepared By** P. Pascal

PHILIPS RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0020.0570 Pos Flow Cal: dP2 at 77,3 (Setpoint 75) [17m 24	ADC Cnts	LE 46633	45703	✓	
37	0020.0580 Pos Flow Cal: dP2 at 87,0 (Setpoint 85) [17m 24	ADC Cnts	LE 47665	46634	✓	
38	0020.0590 Pos Flow Cal: dP2 at 96,5 (Setpoint 95) [17m 24	ADC Cnts	LE 48466	47666	✓	
39	0020.0600 Pos Flow Cal: dP2 at 107,2 (Setpoint 105) [17m	ADC Cnts	LE 49350	48468	✓	
40	0020.0610 Pos Flow Cal: dP2 at 118,3 (Setpoint 115) [17m	ADC Cnts	LE 49981	49351	✓	
41	0020.0620 Pos Flow Cal: dP2 at 127,7 (Setpoint 125) [17m	ADC Cnts	LE 50762	49982	✓	
42	0020.0630 Pos Flow Cal: dP2 at 138,7 (Setpoint 135) [17m	ADC Cnts	LE 51310	50763	✓	
43	0020.0640 Pos Flow Cal: dP2 at 147,3 (Setpoint 145) [17m	ADC Cnts	LE 52629	51311	✓	
44	0020.0650 Pos Flow Cal: dP2 at 168,5 (Setpoint 170) [17m	ADC Cnts	LE 53744	52630	✓	
45	0020.0660 Pos Flow Cal: dP2 at 186,7 (Setpoint 190) [17m	ADC Cnts	LE 65435	53745	✓	
46	0020.0900 Atmospheric Press Sensors Cal (Not attempted) [	N/A	Pass	FALSE		✘
47	0020.0920 Write Sensors Cal Table (Not attempted) [17m 24	N/A	Pass	FALSE		✘
48	0020.0970 Write Sensors Cal Table (Not attempted) [17m 24	N/A	Pass	FALSE		✘
49	0020.0935 O2 Sensor Cal (Not attempted) [17m 24s]	N/A	Pass	FALSE		✘
50	0020.0980 dP Calibration: dP1 @ 10,00cmH2O (Setpoint 10,0	ADC Cnts	13061 to 66211	35235	✓	
51	0020.0981 dP Calibration: dP1 @ 4,00cmH2O (Setpoint 4,00)	ADC Cnts	194661 to 230511	208303	✓	
52	0020.0982 dP Calibration: dP1 @ 0,00cmH2O (Setpoint 0,00)	ADC Cnts	297270 to 358070	327891	✓	
53	0020.0983 dP Calibration: dP1 @ 4,00cmH2O (Setpoint 4,00)	ADC Cnts	425271 to 461121	447519	✓	
54	0020.0984 dP Calibration: dP1 @ 10,00cmH2O (Setpoint 10,0	ADC Cnts	589571 to 642721	619915	✓	
55	0020.1020 Write dP Cal Table [21m 59s]	N/A	Pass	TRUE	✓	
56	0020.1030 Write Image Table [22m 12s]	N/A	Pass	TRUE	✓	

Test Started On 03/26/24 07:44:26

Serial Number TV016072711 **Elapsed Test Time** 22m 12s

Status **ERROR**

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Troubleshooting: Group Level mode.
File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 77D6, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20