Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0020.0570 Pos Flow Cal: dP2 at 77,2 (Setpoint 75) [12m 53	ADC Cnts	LE 47776	46698	✓	
37	0020.0580 Pos Flow Cal: dP2 at 88,3 (Setpoint 85) [12m 53	ADC Cnts	LE 48398	47777	✓	
38	0020.0590 Pos Flow Cal: dP2 at 95,5 (Setpoint 95) [12m 53	ADC Cnts	LE 49199	48399	✓	
39	0020.0600 Pos Flow Cal: dP2 at 105,0 (Setpoint 105) [12m	ADC Cnts	LE 49940	49200	✓	
40	0020.0610 Pos Flow Cal: dP2 at 114,7 (Setpoint 115) [12m	ADC Cnts	LE 50734	49941	✓	
41	0020.0620 Pos Flow Cal: dP2 at 127,3 (Setpoint 125) [12m	ADC Cnts	LE 51384	50735	✓	
42	0020.0630 Pos Flow Cal: dP2 at 136,7 (Setpoint 135) [12m	ADC Cnts	LE 51995	51385	✓	
43	0020.0640 Pos Flow Cal: dP2 at 146,3 (Setpoint 145) [12m	ADC Cnts	LE 53251	51996	✓	
44	0020.0650 Pos Flow Cal: dP2 at 168,7 (Setpoint 170) [12m	ADC Cnts	LE 54447	53252	✓	
45	0020.0660 Pos Flow Cal: dP2 at 188,8 (Setpoint 190) [12m	ADC Cnts	LE 65435	54448	✓	
46	0020.0900 Atmospheric Press Sensors Cal Patm: Slope (m) [	nHg/ADC	EQ 0,000517	0,000517	✓	
47	0020.0910 Atmospheric Press Sensors Cal Patm: Intercept (	inHg	1,91 to 4,09	2,67	✓	
48	0020.0920 Write Sensors Cal Table [13m 18s]	N/A	Pass	TRUE	✓	
49	0020.0970 Write Pprox Cal Table [13m 37s]	N/A	Pass	TRUE	✓	
50	0020.0935 O2 Sensor Cal (not attempted for SW ver.14.2.05	N/A	None	None	✓	
51	0020.0980 dP Calibration: dP1 @ 10,00cmH2O (Setpoint 10,0	ADC Cnts	11494 to 64644	34604	✓	
52	0020.0981 dP Calibration: dP1 @ 4,00cmH2O (Setpoint 4,00)	ADC Cnts	193094 to 228944	205316	✓	
53	0020.0982 dP Calibration: dP1 @ 0,00cmH2O (Setpoint 0,00)	ADC Cnts	297270 to 358070	326324	✓	
54	0020.0983 dP Calibration: dP1 @ 4,00cmH2O (Setpoint 4,00)	ADC Cnts	423704 to 459554	448326	✓	
55	0020.0984 dP Calibration: dP1 @ 10,00cmH2O (Setpoint 10,0	ADC Cnts	588004 to 641154	624214	✓	
56	0020.1020 Write dP Cal Table [17m 20s]	N/A	Pass	TRUE	✓	
57	0020.1030 Write Image Table [17m 32s]	N/A	Pass	TRUE	✓	

Test Started On 06/03/24 05:20:53

Serial Number TV015081808

Elapsed Test Time 17m 32s

Status **PASS**

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Troubleshooting: Group Level mode.
File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 77D6, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20