Test Software P/N 4102483 T

Tested Part Number LA1054096

Test Software Name Trilogy FSA (PM TEST)



**Software Revision** 24.0.0.0

Prepared By P. Pascal

## Test Report

Gr.   Dissipation			II			Ec.21
Step	Description	Units	Limits	Results	Pass	rail
176	0050.0180 Device's SD Card Inserted: SD Detected (3) [55m	N/A	EQ YES	YES	✓	
177	0050.0190 Device's SD Card Inserted: SD Recognized (3) [5	N/A	EQ YES	YES	✓	
178	0050.0195 Device's SD Card Inserted: SD Unlocked (3) [55m	N/A	EQ YES	YES	✓	
179	0050.0200 Significant Event Log (125 entries): [57m 45s]	N/A	Pass	TRUE	✓	
180	0050.0210 Engineering Debug Log (98 entries): [59m 54s]	N/A	Pass	TRUE	✓	
181	0050.0020 Reset Operational Hours (4690,5 Hrs) [59m 54s]	N/A	Pass	TRUE	✓	
						-
						-
						-
						_
						_
						-
						_
						-
						I
						<u> </u>
						<u> </u>
					-	
					$\perp$	

**Test Started On** 01/04/24 02:55:24

Serial Number TV11909263D Elapsed Test Time 59m 54s

 $^{Status}$  PASS

**Page** 6 **of** 6