Test Software P/N 4102483

**Tested Part Number** 1040000

Test Software Name Trilogy FSA (REPAIR TEST)



**Software Revision** 24.0.0.0

Prepared By P. Pascal

## Test Report

Tost Report					
Step	Description	Units	Limits	Results	Pass Fail
36	0021.0570 02 Positive Flow Cal: dP1 at 64,7 (SP 65 @80,9P	ADC Cnts	LE 2016	1873	<u> </u>
37	0021.0580 02 Positive Flow Cal: dP1 at 74,0 (SP 75 @80,0P	ADC Cnts	LE 2149	2017	<b>√</b>
38	0021.0590 O2 Positive Flow Cal: dP1 at 83,7 (SP 85 @79,2P	ADC Cnts	LE 2289	2150	<b>√</b>
39	0021.0600 O2 Positive Flow Cal: dP1 at 95,0 (SP 95 @78,3P	ADC Cnts	LE 2417	2290	<b>√</b>
40	0021.0610 O2 Positive Flow Cal: dP1 at 104,3 (SP 105 @77,	ADC Cnts	LE 2536	2418	<b>√</b>
41	0021.0620 O2 Positive Flow Cal: dP1 at 114,2 (SP 115 @76,	ADC Cnts	LE 2641	2537	<b>√</b>
42	0021.0630 O2 Positive Flow Cal: dP1 at 123,8 (SP 125 @75,	ADC Cnts	LE 2747	2642	<b>√</b>
43	0021.0640 O2 Positive Flow Cal: dP1 at 135,0 (SP 135 @74,	ADC Cnts	LE 2835	2748	✓
44	0021.0650 O2 Positive Flow Cal: dP1 at 144,3 (SP 145 @73,	ADC Cnts	LE 3082	2836	<b>√</b>
45	0021.0660 O2 Positive Flow Cal: dP1 at 174,0 (SP 175 @70,	ADC Cnts	LE 3209	3083	<b>√</b>
46	0021.0670 O2 Positive Flow Cal: dP1 at 190,2 (SP 190 @68,	ADC Cnts	LE 3999	3210	✓
47	0021.0880 02 Sensor Cal (NO_O2_ERROR): 536 mV at ambient	N/A	Pass	TRUE	✓
48	0021.0930 Write OBM Sensors Cal Table (V: TRUE, A: TRUE)	N/A	Pass	TRUE	✓

**Test Started On** 02/20/24 11:51:04

Serial Number TV012043013 Elapsed Test Time 15m 14s

Status PASS

Page 2 **of** 2