Test Software P/N 4102483

Tested Part Number 1040000

Test Software Name Trilogy FSA (REPAIR TEST)



Software Revision 24.0.0.0

Prepared By P. Pascal

Test Report

Step	Description	Units	Limits	Results	Pass Fail
36	0021.0570 O2 Positive Flow Cal: dP1 at 64,3 (SP 65 @76,5P)	ADC Cnts	LE 1859	1716	√
37	0021.0580 02 Positive Flow Cal: dP1 at 74,2 (SP 75 @75,6P)	ADC Cnts	LE 1993	1860	✓
38	0021.0590 O2 Positive Flow Cal: dP1 at 84,8 (SP 85 @74,9P:	ADC Cnts	LE 2117	1994	√
39	0021.0600 O2 Positive Flow Cal: dP1 at 94,2 (SP 95 @74,1P:	ADC Cnts	LE 2251	2118	√
40	0021.0610 O2 Positive Flow Cal: dP1 at 104,5 (SP 105 @73,	ADC Cnts	LE 2368	2252	√
41	0021.0620 O2 Positive Flow Cal: dP1 at 114,8 (SP 115 @72,	ADC Cnts	LE 2475	2369	√
42	0021.0630 02 Positive Flow Cal: dP1 at 124,5 (SP 125 @71,	ADC Cnts	LE 2581	2476	√
43	0021.0640 O2 Positive Flow Cal: dP1 at 134,0 (SP 135 @70,	ADC Cnts	LE 2671	2582	√
44	0021.0650 O2 Positive Flow Cal: dP1 at 144,0 (SP 145 @69,	ADC Cnts	LE 2935	2672	
45	0021.0660 O2 Positive Flow Cal: dP1 at 174,2 (SP 175 @66,	ADC Cnts	LE 3061	2936	→
46	0021.0670 O2 Positive Flow Cal: dP1 at 189,8 (SP 190 @64,	ADC Cnts	LE 3999	3062	
47	0021.0880 02 Sensor Cal (NO_O2_ERROR): 605 mV at ambient a	N/A	Pass	TRUE	→
48	0021.0930 Write OBM Sensors Cal Table (V: TRUE, A: TRUE)	N/A	Pass	TRUE	→

Test Started On 04/23/24 01:41:10

Elapsed Test Time 14m 17s

Status PASS

Serial Number TV02001060E

Page 2 **of** 2