Test Software P/N 4102483 **Tested Part Number** LA1054096

Test Software Name Trilogy FSA (PM TEST)

Software Revision 24.0.0.0 **Prepared By** P. Pascal

PHILIPS RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
71	0030.1030 Contr.Press.Ver.(35876): P1 at 46,32 (SP 46, 1,	cmH2O	45,32 to 47,32	45,99	✓	
72	0030.1040 Monit.Press.Ver.(36198): P2 at 46,32 (SP 46, 1,	cmH2O	45,32 to 47,32	46,07	✓	
73	0030.0164 Pprox Press. Ver.(35834): P2 at 46,32 (SP 46, 1	cmH2O	45,32 to 47,32	46,08	✓	
74	0030.1050 Contr.Press.Ver.(38320): P1 at 50,34 (SP 50, 1,	cmH2O	49,34 to 51,34	50,00	✓	
75	0030.1060 Monit.Press.Ver.(38987): P2 at 50,34 (SP 50, 1,	cmH2O	49,34 to 51,34	50,06	✓	
76	0030.0165 Pprox Press. Ver.(38196): P2 at 50,34 (SP 50, 1	cmH2O	49,34 to 51,34	50,06	✓	
77	0030.1070 Contr.Press.Ver.(46008): P1 at 63,12 (SP 65, 1,	cmH2O	62,12 to 64,12	62,65	✓	
78	0030.1080 Monit.Press.Ver.(47870): P2 at 63,12 (SP 65, 1,	cmH2O	62,12 to 64,12	62,76	✓	
79	0030.0166 Pprox Press. Ver.(45922): P2 at 63,12 (SP 65, 1	cmH2O	62,12 to 64,12	62,88	✓	
80	0030.1150 Atmospheric Pressure Verify: Patm at 28,05,28,0	inHg	27,45 to 28,65	28,45	✓	
81	0030.1250 Airstream Temperature (Tair) at 24,00 [18m 21s]	C deg	19,00 to 39,00	27,14	✓	

Test Started On 04/26/24 08:58:50

Serial Number TV119091027 **Elapsed Test Time** 18m 21s

Status **FAIL**

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Production mode.
File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20