Test Software P/N 4102483

Tested Part Number 1040000

Test Software Name Trilogy FSA (REPAIR TEST)



Software Revision 24.0.0.0

Prepared By P. Pascal

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0021.0570 O2 Positive Flow Cal: dP1 at 64,5 (SP 65 @78,6P	ADC Cnts	LE 1919	1777	√	一
37	0021.0580 O2 Positive Flow Cal: dP1 at 74,5 (SP 75 @76,7P	ADC Cnts	LE 2053	1920	√	
38	0021.0590 02 Positive Flow Cal: dP1 at 84,0 (SP 85 @75,0P	ADC Cnts	LE 2192	2054	√	
39	0021.0600 O2 Positive Flow Cal: dP1 at 95,3 (SP 95 @73,3P	ADC Cnts	LE 2296	2193	√	
40	0021.0610 02 Positive Flow Cal: dP1 at 104,8 (SP 105 @72,	ADC Cnts	LE 2409	2297	√	
41	0021.0620 02 Positive Flow Cal: dP1 at 114,8 (SP 115 @70,	ADC Cnts	LE 2513	2410	√	
42	0021.0630 O2 Positive Flow Cal: dP1 at 124,3 (SP 125 @69,	ADC Cnts	LE 2615	2514	√	
43	0021.0640 O2 Positive Flow Cal: dP1 at 134,7 (SP 135 @68,	ADC Cnts	LE 2702	2616	√	
44	0021.0650 O2 Positive Flow Cal: dP1 at 145,0 (SP 145 @67,	ADC Cnts	LE 2966	2703	√	
45	0021.0660 O2 Positive Flow Cal: dP1 at 174,3 (SP 175 @62,	ADC Cnts	LE 3100	2967	√	
46	0021.0670 O2 Positive Flow Cal: dP1 at 188,7 (SP 190 @59,	ADC Cnts	LE 3999	3101	√	
47	0021.0880 O2 Sensor Cal (RANGE_FAILURE): 3 mV at ambient	N/A	Pass	FALSE		×
48	0021.0930 Write OBM Sensors Cal Table [22m 29s]	N/A	Pass	FALSE		×
						\vdash
						\vdash
					+	\vdash
						\vdash
						\vdash
					+	
					+	\vdash
					-	├
					-	├─
						├─
					_	├
						—
						Ь—
						<u> </u>
						<u> </u>
						<u> </u>

Test Started On 02/05/24 09:17:57

Elapsed Test Time 22m 29s

Serial Number TV014061613

Status FAIL

Page 2 **of** 2