Test Software P/N 4102483

Tested Part Number LA1054096

Test Software Name Trilogy FSA (PM TEST)

Software Revision 24.0.0.0

Prepared By P. Pascal

PHILIPS

RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
176	0050.0180 Device's SD Card Inserted: SD Detected (3) [59m	N/A	EQ YES	YES	✓	
177	0050.0190 Device's SD Card Inserted: SD Recognized (3) [5	N/A	EQ YES	YES	✓	
178	0050.0195 Device's SD Card Inserted: SD Unlocked (3) [59m	N/A	EQ YES	YES	✓	
179	0050.0200 Significant Event Log (123 entries): [1 hr 2m 2	N/A	Pass	TRUE	✓	
180	0050.0210 Engineering Debug Log (121 entries): [1 hr 3m	N/A	Pass	TRUE	✓	
181	0050.0020 Reset Operational Hours (0,0 Hrs) [1 hr 3m 46s]	N/A	Pass	TRUE	✓	

Test Started On 01/12/24 02:34:57

Serial Number TV119042613

Elapsed Test Time 1 hr 3m 46s

Status **PASS**