

Test Software P/N 4102483Tested Part Number LA1054096Test Software Name Trilogy FSA (PM TEST)Software Revision 24.0.0.0Prepared By P. Pascal

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0030.0280 Pos Flow Verify: dP2 at 47,5 (Setpoint 50) [9m	SLPM	42,1 to 52,9	48,2	✓	
37	0030.0290 Pos Flow Verify: dP2 at 37,0 (Setpoint 40) [9m	SLPM	32,0 to 42,0	37,6	✓	
38	0030.0300 Pos Flow Verify: dP2 at 28,0 (Setpoint 30) [9m	SLPM	23,4 to 32,6	27,6	✓	
39	0030.0310 Pos Flow Verify: dP2 at 19,3 (Setpoint 20) [9m	SLPM	14,3 to 24,3	19,1	✓	
40	0030.0320 Pos Flow Verify: dP2 at 10,0 (Setpoint 10) [9m	SLPM	5,5 to 14,5	10,0	✓	
41	0030.0330 Pos Flow Verify: dP2 at 4,8 (Setpoint 5) [9m 22	SLPM	1,1 to 8,6	5,0	✓	
42	0030.0510 Zero Flow Verify: dP2 at 0,0 (Setpoint 0) [9m 3	SLPM	-2,0 to 2,0	0,0	✓	
43	0030.0550 Neg Flow Verify: dP2 at -138,8 (Setpoint 140) [SLPM	-147,9 to -129,8	-140,2	✓	
44	0030.0560 Neg Flow Verify: dP2 at -128,2 (Setpoint 130) [SLPM	-136,8 to -119,5	-129,2	✓	
45	0030.0570 Neg Flow Verify: dP2 at -117,8 (Setpoint 120) [SLPM	-126,0 to -109,6	-119,2	✓	
46	0030.0580 Neg Flow Verify: dP2 at -108,0 (Setpoint 110) [SLPM	-115,8 to -100,2	-109,6	✓	
47	0030.0590 Neg Flow Verify: dP2 at -98,0 (Setpoint 100) [1	SLPM	-105,4 to -90,6	-99,3	✓	
48	0030.0600 Neg Flow Verify: dP2 at -88,5 (Setpoint 90) [13	SLPM	-95,5 to -81,5	-89,1	✓	
49	0030.0610 Neg Flow Verify: dP2 at -77,8 (Setpoint 80) [13	SLPM	-84,4 to -71,2	-78,2	✓	
50	0030.0620 Neg Flow Verify: dP2 at -67,5 (Setpoint 70) [13	SLPM	-73,7 to -61,3	-68,5	✓	
51	0030.0630 Neg Flow Verify: dP2 at -58,0 (Setpoint 60) [13	SLPM	-63,8 to -52,2	-58,9	✓	
52	0030.0640 Neg Flow Verify: dP2 at -48,0 (Setpoint 50) [13	SLPM	-53,4 to -42,6	-48,2	✓	
53	0030.0650 Neg Flow Verify: dP2 at -38,0 (Setpoint 40) [13	SLPM	-43,0 to -33,0	-38,4	✓	
54	0030.0660 Neg Flow Verify: dP2 at -28,0 (Setpoint 30) [13	SLPM	-32,6 to -23,4	-28,3	✓	
55	0030.0670 Neg Flow Verify: dP2 at -19,7 (Setpoint 20) [13	SLPM	-24,7 to -14,7	-19,4	✓	
56	0030.0680 Neg Flow Verify: dP2 at -9,3 (Setpoint 10) [13m	SLPM	-13,8 to -4,8	-9,5	✓	
57	0030.0690 Neg Flow Verify: dP2 at -5,0 (Setpoint 5) [13m	SLPM	-8,8 to -1,2	-5,0	✓	
58	0030.0700 Neg Flow Verify: dP2 at -0,0 (Setpoint 0) [13m	SLPM	-2,0 to 2,0	0,0	✓	
59	0030.0950 Contr.Press.Ver.(8794): P1 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,00	✓	
60	0030.0960 Monit.Press.Ver.(4109): P2 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,00	✓	
61	0030.0160 Pprox Press. Ver.(8866): P2 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,00	✓	
62	0030.0970 Contr.Press.Ver.(11205): P1 at 3,99 (SP 4, 1, 1)	cmH2O	2,99 to 4,99	3,99	✓	
63	0030.0980 Monit.Press.Ver.(6875): P2 at 3,99 (SP 4, 1, 1)	cmH2O	2,99 to 4,99	3,95	✓	
64	0030.0161 Pprox Press. Ver.(11275): P2 at 3,99 (SP 4, 1,	cmH2O	2,99 to 4,99	4,01	✓	
65	0030.0990 Contr.Press.Ver.(19613): P1 at 17,99 (SP 18, 1,	cmH2O	16,99 to 18,99	17,99	✓	
66	0030.1000 Monit.Press.Ver.(16536): P2 at 17,99 (SP 18, 1,	cmH2O	16,99 to 18,99	17,91	✓	
67	0030.0162 Pprox Press. Ver.(19664): P2 at 17,99 (SP 18, 1	cmH2O	16,99 to 18,99	18,05	✓	
68	0030.1010 Contr.Press.Ver.(28013): P1 at 31,99 (SP 32, 1,	cmH2O	30,99 to 32,99	31,99	✓	
69	0030.1020 Monit.Press.Ver.(26220): P2 at 31,99 (SP 32, 1,	cmH2O	30,99 to 32,99	31,90	✓	
70	0030.0163 Pprox Press. Ver.(27998): P2 at 31,99 (SP 32, 1	cmH2O	30,99 to 32,99	31,97	✓	

Test Started On 02/16/24 12:49:01Serial Number TV114121231Elapsed Test Time 31m 59sStatus **TERMINATED**

Page 2 of 5

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Production mode.
 File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
 Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20