Test Software P/N 4102483 **Tested Part Number** 1040000

Test Software Name Trilogy FSA (REPAIR TEST)

Software Revision 24.0.0.0 **Prepared By** P. Pascal

PHILIPS RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0020.0570 Pos Flow Cal: dP2 at 77,3 (Setpoint 75) [18m 22	ADC Cnts	LE 46050	45226	✓	
37	0020.0580 Pos Flow Cal: dP2 at 85,7 (Setpoint 85) [18m 22	ADC Cnts	LE 47231	46051	✓	
38	0020.0590 Pos Flow Cal: dP2 at 96,8 (Setpoint 95) [18m 22	ADC Cnts	LE 48216	47232	✓	
39	0020.0600 Pos Flow Cal: dP2 at 109,2 (Setpoint 105) [18m	ADC Cnts	LE 48807	48217	✓	
40	0020.0610 Pos Flow Cal: dP2 at 116,3 (Setpoint 115) [18m	ADC Cnts	LE 49916	48808	✓	
41	0020.0620 Pos Flow Cal: dP2 at 132,3 (Setpoint 125) [18m	ADC Cnts	LE 50141	49917	✓	
42	0020.0630 Pos Flow Cal: dP2 at 135,3 (Setpoint 135) [18m	ADC Cnts	LE 50798	50142	✓	
43	0020.0640 Pos Flow Cal: dP2 at 146,7 (Setpoint 145) [18m	ADC Cnts	LE 52153	50799	✓	
44	0020.0650 Pos Flow Cal: dP2 at 169,5 (Setpoint 170) [18m	ADC Cnts	LE 53242	52154	✓	
45	0020.0660 Pos Flow Cal: dP2 at 188,0 (Setpoint 190) [18m	ADC Cnts	LE 65435	53242	✓	
46	0020.0900 Atmospheric Press Sensors Cal Patm: Slope (m) [	inHg/ADC	EQ 0,000517	0,000517	✓	
47	0020.0910 Atmospheric Press Sensors Cal Patm: Intercept (	inHg	1,91 to 4,09	2,66	✓	
48	0020.0920 Write Sensors Cal Table [18m 48s]	N/A	Pass	TRUE	✓	
49	0020.0970 Write Pprox Cal Table [19m 4s]	N/A	Pass	TRUE	✓	
50	0020.0935 O2 Sensor Cal (not attempted for SW ver.14.2.05	N/A	None	None	✓	
51	0020.0980 dP Calibration: dP1 @ 10,00cmH2O (Setpoint 10,0	ADC Cnts	14554 to 67704	45993	✓	
52	0020.0981 dP Calibration: dP1 @ 4,00cmH2O (Setpoint 4,00)	ADC Cnts	196154 to 232004	212940	✓	
53	0020.0982 dP Calibration: dP1 @ 0,00cmH2O (Setpoint 0,00)	ADC Cnts	297270 to 358070	329384	✓	
54	0020.0983 dP Calibration: dP1 @ 4,00cmH2O (Setpoint 4,00)	ADC Cnts	426764 to 462614	446782	✓	
55	0020.0984 dP Calibration: dP1 @ 10,00cmH2O (Setpoint 10,0	ADC Cnts	591064 to 644214	618719	✓	
56	0020.1020 Write dP Cal Table [27m 7s]	N/A	Pass	TRUE	✓	
57	0020.1030 Write Image Table [27m 23s]	N/A	Pass	TRUE	✓	

Test Started On 02/06/24 10:54:36

Serial Number TV02001060E **Elapsed Test Time** 27m 23s

Status **PASS**

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Troubleshooting: Group Level mode.
File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 77D6, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20