Test Software P/N 4102483

**Tested Part Number** 1040000

Test Software Name Trilogy FSA (REPAIR TEST)



**Software Revision** 24.0.0.0

Prepared By P. Pascal

## Test Report

Step	Description	Units	Limits	Results	Pass Fail
36	0021.0570 O2 Positive Flow Cal: dPl at 64,8 (SP 65 @81,4P)	ADC Cnts	LE 1919	1781	<b>√</b>
37	0021.0580 O2 Positive Flow Cal: dP1 at 74,3 (SP 75 @80,0P)	ADC Cnts	LE 2052	1920	<b>√</b>
38	0021.0590 O2 Positive Flow Cal: dP1 at 84,2 (SP 85 @78,6P)	ADC Cnts	LE 2183	2053	<b> </b>
39	0021.0600 O2 Positive Flow Cal: dP1 at 94,2 (SP 95 @76,9P	ADC Cnts	LE 2315	2184	<b>√</b>
40	0021.0610 O2 Positive Flow Cal: dP1 at 105,2 (SP 105 @75,	ADC Cnts	LE 2425	2316	<b> </b>
41	0021.0620 02 Positive Flow Cal: dP1 at 114,7 (SP 115 @73,	ADC Cnts	LE 2526	2426	<b> </b>
42	0021.0630 O2 Positive Flow Cal: dP1 at 124,5 (SP 125 @71,	ADC Cnts	LE 2623	2527	<b>√</b>
43	0021.0640 02 Positive Flow Cal: dP1 at 133,9 (SP 135 @70,	ADC Cnts	LE 2714	2624	<b>√</b>
44	0021.0650 O2 Positive Flow Cal: dP1 at 145,2 (SP 145 @68,	ADC Cnts	LE 2981	2715	<b> </b>
45	0021.0660 O2 Positive Flow Cal: dP1 at 174,2 (SP 175 @65,	ADC Cnts	LE 3114	2982	<b> </b>
46	0021.0670 O2 Positive Flow Cal: dP1 at 190,1 (SP 190 @62,	ADC Cnts	LE 3999	3115	<b> </b>
47	0021.0880 02 Sensor Cal (NO_O2_ERROR): 537 mV at ambient a	N/A	Pass	TRUE	<b>│</b>
48	0021.0930 Write OBM Sensors Cal Table (V: TRUE, A: TRUE)	N/A	Pass	TRUE	<b> </b>

**Test Started On** 06/05/24 01:13:49

Serial Number TV019020104 Elapsed Test Time 21m 33s

Status PASS

Page 2 **of** 2