Test Software P/N 4102483

Tested Part Number 1040000

Test Software Name Trilogy FSA (REPAIR TEST)



Software Revision 24.0.0.0

Prepared By P. Pascal

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0021.0570 O2 Positive Flow Cal: dP1 at 64,0 (SP 65 @78,4P)	ADC Cnts	LE 1858	1713	√	
37	0021.0580 O2 Positive Flow Cal: dP1 at 74,7 (SP 75 @77,9P	ADC Cnts	LE 1988	1859	√	
38	0021.0590 O2 Positive Flow Cal: dP1 at 84,2 (SP 85 @77,4P	ADC Cnts	LE 2114	1989	√	
39	0021.0600 02 Positive Flow Cal: dP1 at 94,6 (SP 95 @76,7P	ADC Cnts	LE 2243	2115	√	
40	0021.0610 02 Positive Flow Cal: dP1 at 105,2 (SP 105 @76,	ADC Cnts	LE 2338	2244	√	
41	0021.0620 O2 Positive Flow Cal: dPl at 113,7 (SP 115 @75,	ADC Cnts	LE 2447	2339	√	
42	0021.0630 O2 Positive Flow Cal: dPl at 124,7 (SP 125 @74,	ADC Cnts	LE 2539	2448	√	
43	0021.0640 O2 Positive Flow Cal: dP1 at 133,7 (SP 135 @73,	ADC Cnts	LE 2624	2540	√	
44	0021.0650 O2 Positive Flow Cal: dP1 at 143,3 (SP 145 @73,	ADC Cnts	LE 2884	2624	√	
45	0021.0660 O2 Positive Flow Cal: dP1 at 174,0 (SP 175 @70,	ADC Cnts	LE 3033	2885	√	
46	0021.0670 O2 Positive Flow Cal: dP1 at 191,2 (SP 190 @68,	ADC Cnts	LE 3999	3034	√	
47	0021.0880 02 Sensor Cal (RANGE_FAILURE): 9 mV at ambient	N/A	Pass	FALSE		×
48	0021.0930 Write OBM Sensors Cal Table [18m 36s]	N/A	Pass	FALSE		×
						$\overline{}$
						-
						\vdash
						<u> </u>
						
						\vdash
					_	
						<u> </u>
						<u> </u>
						<u> </u>
						<u> </u>
						<u> </u>
						<u> </u>
						l

Test Started On 08/30/24 10:34:55

Serial Number TV019111110 Elapsed Test Time 18m 36s

Status FAIL

Page 2 **of** 2