Test Software P/N 4102483

Tested Part Number LA1054096

 $\textbf{Test Software Name} \ \underline{\texttt{Trilogy FSA}} \ (\texttt{PM TEST})$



Software Revision 24.0.0.0

Prepared By P. Pascal

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
71	0030.1030 Contr.Press.Ver.(35797): P1 at 44,60 (SP 46, 1,	cmH20	43,60 to 45,60	46,00		×
72	0030.1040 Monit.Press.Ver.(34421): P2 at 44,60 (SP 46, 1,	cmH20	43,60 to 45,60	44,50	√	
73	0030.0164 Pprox Press. Ver.(8928): P2 at 44,60 (SP 46, 1,	cmH20	43,60 to 45,60	0,01		×
74	0030.1050 Contr.Press.Ver.(38206): P1 at 48,65 (SP 50, 1,	cmH20	47,65 to 49,65	49,99		×
75	0030.1060 Monit.Press.Ver.(37235): P2 at 48,65 (SP 50, 1,	cmH20	47,65 to 49,65	48,54	√	
76	0030.0165 Pprox Press. Ver.(8926): P2 at 48,65 (SP 50, 1,	cmH2O	47,65 to 49,65	0,01		×
77	0030.1070 Contr.Press.Ver.(46823): P1 at 63,01 (SP 65, 1,	cmH20	62,01 to 64,01	64,30		×
78	0030.1080 Monit.Press.Ver.(47202): P2 at 63,01 (SP 65, 1,	cmH20	62,01 to 64,01	62,85	√	
79	0030.0166 Pprox Press. Ver.(8922): P2 at 63,01 (SP 65, 1,	cmH20	62,01 to 64,01	0,00		×
8 0	0030.1150 Atmospheric Pressure Verify: Patm at 28,35,28,3	inHg	27,75 to 28,95	28,65	√	
81	0030.1250 Airstream Temperature (Tair) at 23,00 [22m 45s]	C deg	18,00 to 38,00	26,89	✓	
						<u> </u>
						
						<u> </u>
						<u> </u>
						-

Test Started On 07/24/24 08:45:34

Serial Number TV115081216 Elapsed Test Time 22m 45s

Status FAIL

Page 3 **of** 3