

Test Software P/N 4102483Tested Part Number 1040000Test Software Name Trilogy FSA (REPAIR TEST)Software Revision 24.0.0.0Prepared By P. Pascal

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
1	0020.0001 Ambient Air: RH=47,40%; T=20,80Cdeg [24s]	N/A	Pass	TRUE	✓	
2	0020.0010 O2 Sensor Heater On (not attempted for SW ver.1)	N/A	None	None	✓	
3	0020.0010 Press Sensors Cal P1@3,99&50,03cmH2O/10047&3338	20/ADC C	0,00151 to 0,0018	0,00167	✓	
4	0020.0020 Press Sensors Cal P1@3,99&50,03cmH2O/10047&3338	cmH2O	-18,22 to -9,05	-13,84	✓	
5	0020.0030 Press Sensors Cal P1@3,99&50,03cmH2O/10047&3338	ADC Cnts	5994 to 9998	8212	✓	
6	0020.0040 Press Sensors Cal P2@3,99&50,03cmH2O/10047&3338	20/ADC C	0,00114 to 0,0018	0,00144	✓	
7	0020.0050 Press Sensors Cal P2@3,99&50,03cmH2O/10047&3338	cmH2O	-19,53 to 0,00	-5,93	✓	
8	0020.0060 Press Sensors Cal P2@3,99&50,03cmH2O/10047&3338	ADC Cnts	0 to 10399	4037	✓	
9	0020.0950 Press Sensors Cal Pprox@3,99&50,03cmH2O/10047&3	20/ADC C	0,00151 to 0,0018	0,00167	✓	
10	0020.0960 Press Sensors Cal Pprox@3,99&50,03cmH2O/10047&3	cmH2O	-18,22 to -9,05	-13,61	✓	
11	0020.0970 Press Sensors Cal Pprox@3,99&50,03cmH2O/10047&3	ADC Cnts	5994 to 9998	8085	✓	
12	0020.0120 Neg Flow Cal: dP2 at -146,0 (Setpoint 145) [10m	ADC Cnts	GE 101	14205	✓	
13	0020.0130 Neg Flow Cal: dP2 at -132,9 (Setpoint 135) [10m	ADC Cnts	GE 14206	15076	✓	
14	0020.0140 Neg Flow Cal: dP2 at -123,2 (Setpoint 125) [10m	ADC Cnts	GE 15077	15766	✓	
15	0020.0150 Neg Flow Cal: dP2 at -113,3 (Setpoint 115) [10m	ADC Cnts	GE 15767	16582	✓	
16	0020.0160 Neg Flow Cal: dP2 at -103,0 (Setpoint 105) [10m	ADC Cnts	GE 16583	17463	✓	
17	0020.0170 Neg Flow Cal: dP2 at -93,7 (Setpoint 95) [10m 3	ADC Cnts	GE 17464	18478	✓	
18	0020.0180 Neg Flow Cal: dP2 at -83,5 (Setpoint 85) [10m 3	ADC Cnts	GE 18480	19587	✓	
19	0020.0190 Neg Flow Cal: dP2 at -73,0 (Setpoint 75) [10m 3	ADC Cnts	GE 19588	20697	✓	
20	0020.0200 Neg Flow Cal: dP2 at -64,0 (Setpoint 65) [10m 3	ADC Cnts	GE 20698	21750	✓	
21	0020.0210 Neg Flow Cal: dP2 at -53,0 (Setpoint 55) [10m 3	ADC Cnts	GE 21750	23052	✓	
22	0020.0220 Neg Flow Cal: dP2 at -43,0 (Setpoint 45) [10m 3	ADC Cnts	GE 23054	24476	✓	
23	0020.0230 Neg Flow Cal: dP2 at -33,0 (Setpoint 35) [10m 3	ADC Cnts	GE 24477	25991	✓	
24	0020.0240 Neg Flow Cal: dP2 at -23,0 (Setpoint 25) [10m 3	ADC Cnts	GE 25992	27682	✓	
25	0020.0250 Neg Flow Cal: dP2 at -14,8 (Setpoint 15) [10m 3	ADC Cnts	GE 27683	29480	✓	
26	0020.0260 Neg Flow Cal: dP2 at -5,0 (Setpoint 5) [10m 35s	ADC Cnts	GE 29482	32163	✓	
27	0020.0270 Neg Flow Cal: dP2 at -0,0 (Setpoint 0) [10m 35s	ADC Cnts	GE 32164	33412	✓	
28	0020.0450 Raw Zero Flow: dP2 at 0,0 (Setpoint 0) [10m 35s	ADC Cnts	GE 0	33412	✓	
29	0020.0500 Pos Flow Cal: dP2 at 4,8 (Setpoint 5) [15m 45s]	ADC Cnts	LE 37634	34781	✓	
30	0020.0510 Pos Flow Cal: dP2 at 14,5 (Setpoint 15) [15m 45s	ADC Cnts	LE 39344	37634	✓	
31	0020.0520 Pos Flow Cal: dP2 at 23,0 (Setpoint 25) [15m 45s	ADC Cnts	LE 41616	39345	✓	
32	0020.0530 Pos Flow Cal: dP2 at 38,0 (Setpoint 35) [15m 45s	ADC Cnts	LE 42702	41616	✓	
33	0020.0540 Pos Flow Cal: dP2 at 47,0 (Setpoint 45) [15m 45s	ADC Cnts	LE 43970	42703	✓	
34	0020.0550 Pos Flow Cal: dP2 at 57,0 (Setpoint 55) [15m 45s	ADC Cnts	LE 45162	43970	✓	
35	0020.0560 Pos Flow Cal: dP2 at 67,7 (Setpoint 65) [15m 45s	ADC Cnts	LE 46133	45163	✓	

Test Started On 06/07/24 09:01:54Serial Number TV019020111Elapsed Test Time 22m 54sStatus **PASS**

Page 1 of 2

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Troubleshooting: Group Level mode.
 File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
 Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 77D6, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20