PHILIPS
RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
176	0050.0180 Device's SD Card Inserted: SD Detected (3) [45m	N/A	EQ YES	YES	✓	
177	0050.0190 Device's SD Card Inserted: SD Recognized (3) [4	N/A	EQ YES	YES	✓	
178	0050.0195 Device's SD Card Inserted: SD Unlocked (3) [45m	N/A	EQ YES	YES	✓	
179	0050.0200 Significant Event Log (143 entries): [47m 21s]	N/A	Pass	TRUE	✓	
180	0050.0210 Engineering Debug Log (102 entries): [48m 58s]	N/A	Pass	TRUE	✓	
181	0050.0020 Reset Operational Hours (5610,8 Hrs) [48m 58s]	N/A	Pass	TRUE	✓	

Test Started On 04/01/24 07:39:12

Serial Number TV120092328

Elapsed Test Time 48m 58s

Status **PASS**