Test Software P/N 4102483 Tested Part Number LA1054096

Test Software Name Trilogy FSA (PM TEST)

Software Revision 24.0.0.0 Prepared By P. Pascal

PHILIPS RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
141	0040.0700 SD Locked Card Inserted: SD Detected (7) [27m 1	N/A	EQ YES	YES	✓	
142	0040.0710 SD Locked Card Inserted: SD Recognized (7) [27m	N/A	EQ YES	YES	✓	
143	0040.0720 SD Locked Card Inserted: Write Protected Mode (	N/A	EQ YES	YES	✓	
144	0040.0730 SD Card Removed: SD Detected (4) [27m 23s]	N/A	EQ NO	NO	✓	
145	0040.0740 SD Card Removed: SD Recognized (4) [27m 23s]	N/A	EQ NO	NO	✓	
146	0040.0750 SD Card Removed: Write Protected Mode (4) [27m	N/A	EQ YES	YES	✓	
147	0040.0780 SD Unlocked Card Inserted: Write Mode Enable (3	N/A	Pass	TRUE	✓	
148	0040.0790 SD: File Write [27m 38s]	N/A	Pass	TRUE	✓	
149	0040.0800 SD: File Length [27m 39s]	N/A	Pass	TRUE	✓	
150	0040.0810 SD: File Read [27m 40s]	N/A	Pass	TRUE	✓	
151	0040.0820 SD: File Erase [27m 41s]	N/A	Pass	TRUE	✓	
152	0040.0840 Nurse Call OFF (NO: 9999,00KOhm, NC: 0,00Ohm) [	N/A	Pass	TRUE	✓	
153	0040.0850 Nurse Call ON (NO: 0,00Ohm, NC: 9999,00KOhm) [2	N/A	Pass	TRUE	✓	
154	0040.0900 P1 with Auto Null Valve Closed at 29,99cmH2O [2	cmH2O	26,99 to 32,99	29,99	✓	
155	0040.0910 P1 with Auto Null Valve Opened at 62,40cmH2O [2	cmH2O	-3,00 to 3,00	0,00	✓	
156	0040.0950 Cooling Fans: Stirring Fan [28m 27s]	N/A	Pass	TRUE	✓	
157	0040.0960 Cooling Fans: Battery Fan [28m 29s]	N/A	Pass	TRUE	✓	
158	0040.0970 Cooling Fans: Exhaust Fan [28m 35s]	N/A	Pass	TRUE	✓	
159	0040.1300 Current Consumption in Sleep Mode (1) [29m 36s]	A	LE 0,20	0,00	✓	
160	0040.1400 O2 Low Flow @ 20,10PSIG [31m 43s]	SLPM	GE 7	11	✓	
161	0040.1410 DSP CRC: JTAG ID Code [32m 6s]	N/A	EQ 01F4001D	01F4001D	✓	
162	0040.1420 DSP CRC: DSP Program CRC [32m 6s]	N/A	EQ FE0F	FE0F	✓	

Test Started On 04/03/24 01:49:07

Serial Number TV113102307 Elapsed Test Time 32m 6s

Status **FAIL**

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Production mode.
File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20