PHILIPS
RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0020.0570 Pos Flow Cal: dP2 at 77,2 (Setpoint 75) [12m 53	ADC Cnts	LE 45884	45280	✓	
37	0020.0580 Pos Flow Cal: dP2 at 84,0 (Setpoint 85) [12m 53	ADC Cnts	LE 47190	45885	✓	
38	0020.0590 Pos Flow Cal: dP2 at 97,8 (Setpoint 95) [12m 53	ADC Cnts	LE 47967	47191	✓	
39	0020.0600 Pos Flow Cal: dP2 at 107,0 (Setpoint 105) [12m	ADC Cnts	LE 48761	47968	✓	
40	0020.0610 Pos Flow Cal: dP2 at 118,2 (Setpoint 115) [12m	ADC Cnts	LE 49452	48762	✓	
41	0020.0620 Pos Flow Cal: dP2 at 128,5 (Setpoint 125) [12m	ADC Cnts	LE 50108	49453	✓	
42	0020.0630 Pos Flow Cal: dP2 at 137,5 (Setpoint 135) [12m	ADC Cnts	LE 50734	50109	✓	
43	0020.0640 Pos Flow Cal: dP2 at 147,2 (Setpoint 145) [12m	ADC Cnts	LE 51984	50735	✓	
44	0020.0650 Pos Flow Cal: dP2 at 167,8 (Setpoint 170) [12m	ADC Cnts	LE 53184	51985	✓	
45	0020.0660 Pos Flow Cal: dP2 at 187,7 (Setpoint 190) [12m	ADC Cnts	LE 65435	53185	✓	
46	0020.0900 Atmospheric Press Sensors Cal Patm: Slope (m) [	inHg/ADC	EQ 0,000517	0,000517	✓	
47	0020.0910 Atmospheric Press Sensors Cal Patm: Intercept (	inHg	1,91 to 4,09	2,66	✓	
48	0020.0920 Write Sensors Cal Table [13m 17s]	N/A	Pass	TRUE	✓	
49	0020.0970 Write Pprox Cal Table [13m 42s]	N/A	Pass	TRUE	✓	
50	0020.0935 O2 Sensor Cal (not attempted for SW ver.14.2.05	N/A	None	None	✓	
51	0020.1030 Write Image Table [14m 3s]	N/A	Pass	TRUE	✓	

Test Started On <u>03/22/24 11:52:07</u>

Serial Number <u>TV116062358</u>   **Elapsed Test Time** <u>14m 3s</u>

Status <u>**PASS**</u>