

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0030.0280 Pos Flow Verify: dP2 at 47,8 (Setpoint 50) [10m	SLPM	42,4 to 53,2	51,2	✓	
37	0030.0290 Pos Flow Verify: dP2 at 37,0 (Setpoint 40) [10m	SLPM	32,0 to 42,0	40,1	✓	
38	0030.0300 Pos Flow Verify: dP2 at 28,0 (Setpoint 30) [10m	SLPM	23,4 to 32,6	29,3	✓	
39	0030.0310 Pos Flow Verify: dP2 at 19,8 (Setpoint 20) [10m	SLPM	14,8 to 24,8	20,5	✓	
40	0030.0320 Pos Flow Verify: dP2 at 9,7 (Setpoint 10) [10m	SLPM	5,2 to 14,2	10,3	✓	
41	0030.0330 Pos Flow Verify: dP2 at 5,0 (Setpoint 5) [10m 2	SLPM	1,2 to 8,8	5,0	✓	
42	0030.0510 Zero Flow Verify: dP2 at 0,0 (Setpoint 0) [10m	SLPM	-2,0 to 2,0	-0,2	✓	
43	0030.0550 Neg Flow Verify: dP2 at -138,5 (Setpoint 140) [SLPM	-147,5 to -129,5	-140,8	✓	
44	0030.0560 Neg Flow Verify: dP2 at -128,5 (Setpoint 130) [SLPM	-137,1 to -119,9	-130,6	✓	
45	0030.0570 Neg Flow Verify: dP2 at -118,0 (Setpoint 120) [SLPM	-126,2 to -109,8	-120,5	✓	
46	0030.0580 Neg Flow Verify: dP2 at -107,8 (Setpoint 110) [SLPM	-115,6 to -100,0	-110,7	✓	
47	0030.0590 Neg Flow Verify: dP2 at -98,2 (Setpoint 100) [1	SLPM	-105,6 to -90,7	-100,2	✓	
48	0030.0600 Neg Flow Verify: dP2 at -88,8 (Setpoint 90) [14	SLPM	-95,9 to -81,8	-89,5	✓	
49	0030.0610 Neg Flow Verify: dP2 at -77,5 (Setpoint 80) [14	SLPM	-84,1 to -70,9	-78,4	✓	
50	0030.0620 Neg Flow Verify: dP2 at -67,8 (Setpoint 70) [14	SLPM	-74,0 to -61,6	-70,1	✓	
51	0030.0630 Neg Flow Verify: dP2 at -58,0 (Setpoint 60) [14	SLPM	-63,8 to -52,2	-60,8	✓	
52	0030.0640 Neg Flow Verify: dP2 at -48,0 (Setpoint 50) [14	SLPM	-53,4 to -42,6	-50,7	✓	
53	0030.0650 Neg Flow Verify: dP2 at -38,0 (Setpoint 40) [14	SLPM	-43,0 to -33,0	-41,2	✓	
54	0030.0660 Neg Flow Verify: dP2 at -28,0 (Setpoint 30) [14	SLPM	-32,6 to -23,4	-30,1	✓	
55	0030.0670 Neg Flow Verify: dP2 at -20,0 (Setpoint 20) [14	SLPM	-25,0 to -15,0	-21,9	✓	
56	0030.0680 Neg Flow Verify: dP2 at -10,0 (Setpoint 10) [14	SLPM	-14,5 to -5,5	-10,8	✓	
57	0030.0690 Neg Flow Verify: dP2 at -4,7 (Setpoint 5) [14m	SLPM	-8,4 to -0,9	-5,1	✓	
58	0030.0700 Neg Flow Verify: dP2 at -0,0 (Setpoint 0) [14m	SLPM	-2,0 to 2,0	-0,2	✓	
59	0030.0950 Contr.Press.Ver.(7998): P1 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,06	✓	
60	0030.0960 Monit.Press.Ver.(4187): P2 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,03	✓	
61	0030.0160 Pprox Press. Ver.(8007): P2 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,00	✓	
62	0030.0970 Contr.Press.Ver.(10385): P1 at 4,02 (SP 4, 1, 1)	cmH2O	3,02 to 5,02	4,00	✓	
63	0030.0980 Monit.Press.Ver.(6956): P2 at 4,02 (SP 4, 1, 1)	cmH2O	3,02 to 5,02	3,98	✓	
64	0030.0161 Pprox Press. Ver.(10400): P2 at 4,02 (SP 4, 1,	cmH2O	3,02 to 5,02	3,78	✓	
65	0030.0990 Contr.Press.Ver.(18890): P1 at 18,10 (SP 18, 1,	cmH2O	17,10 to 19,10	17,99	✓	
66	0030.1000 Monit.Press.Ver.(16701): P2 at 18,10 (SP 18, 1,	cmH2O	17,10 to 19,10	17,95	✓	
67	0030.0162 Pprox Press. Ver.(18861): P2 at 18,10 (SP 18, 1	cmH2O	17,10 to 19,10	17,82	✓	
68	0030.1010 Contr.Press.Ver.(27394): P1 at 32,19 (SP 32, 1,	cmH2O	31,19 to 33,19	31,99	✓	
69	0030.1020 Monit.Press.Ver.(26484): P2 at 32,19 (SP 32, 1,	cmH2O	31,19 to 33,19	31,92	✓	
70	0030.0163 Pprox Press. Ver.(27352): P2 at 32,19 (SP 32, 1	cmH2O	31,19 to 33,19	31,86	✓	

Test Started On 04/26/24 09:19:01Serial Number TV119091027Elapsed Test Time 46m 31sStatus **PASS**

Page 2 of 6

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Production mode.
 File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
 Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20