Test Software P/N 4102483

Tested Part Number LA1054096

 $\textbf{Test Software Name} \ \underline{\texttt{Trilogy FSA}} \ (\texttt{PM TEST})$



PHILIPS RESPIRONICS

Software Revision 24.0.0.0

Prepared By P. Pascal

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0030.0280 Pos Flow Verify: dP2 at 47,0 (Setpoint 50) [9m	SLPM	41,6 to 52,4	47,1		_
37	0030.0290 Pos Flow Verify: dP2 at 38,0 (Setpoint 40) [9m	SLPM	33,0 to 43,0	37,9	√	
38	0030.0300 Pos Flow Verify: dP2 at 27,8 (Setpoint 30) [9m	SLPM	23,2 to 32,4	28,2	√	
39	0030.0310 Pos Flow Verify: dP2 at 20,0 (Setpoint 20) [9m	SLPM	15,0 to 25,0	19,9	√	
40	0030.0320 Pos Flow Verify: dP2 at 10,0 (Setpoint 10) [9m	SLPM	5,5 to 14,5	10,0	√	
41	0030.0330 Pos Flow Verify: dP2 at 5,0 (Setpoint 5) [9m 30	SLPM	1,2 to 8,8	4,8	√	
42	0030.0510 Zero Flow Verify: dP2 at 0,0 (Setpoint 0) [9m 3	SLPM	-2,0 to 2,0	0,0	√	
43	0030.0550 Neg Flow Verify: dP2 at -138,5 (Setpoint 140) [SLPM	-147,5 to -129,5	-138,5	✓	
44	0030.0560 Neg Flow Verify: dP2 at -128,0 (Setpoint 130) [SLPM	-136,6 to -119,4	-127,3	√	
45	0030.0570 Neg Flow Verify: dP2 at -117,8 (Setpoint 120) [SLPM	-126,0 to -109,6	-117,2	✓	
46	0030.0580 Neg Flow Verify: dP2 at -108,2 (Setpoint 110) [SLPM	-116,0 to -100,3	-107,7	✓	
47	0030.0590 Neg Flow Verify: dP2 at -98,7 (Setpoint 100) [1	SLPM	-106,1 to -91,2	-98,2	√	
48	0030.0600 Neg Flow Verify: dP2 at -88,7 (Setpoint 90) [14	SLPM	-95,7 to -81,6	-88,0	√	
49	0030.0610 Neg Flow Verify: dP2 at -78,0 (Setpoint 80) [14	SLPM	-84,6 to -71,4	-77,3	✓	
50	0030.0620 Neg Flow Verify: dP2 at -67,2 (Setpoint 70) [14	SLPM	-73,4 to -61,0	-68,1	√	
51	0030.0630 Neg Flow Verify: dP2 at -58,0 (Setpoint 60) [14	SLPM	-63,8 to -52,2	-58,5	✓	
52	0030.0640 Neg Flow Verify: dP2 at -48,0 (Setpoint 50) [14	SLPM	-53,4 to -42,6	-48,1	✓	
53	0030.0650 Neg Flow Verify: dP2 at -38,0 (Setpoint 40) [14	SLPM	-43,0 to -33,0	-38,0	√	
54	0030.0660 Neg Flow Verify: dP2 at -28,0 (Setpoint 30) [14	SLPM	-32,6 to -23,4	-28,3	✓	
55	0030.0670 Neg Flow Verify: dP2 at -20,0 (Setpoint 20) [14	SLPM	-25,0 to -15,0	-20,1	✓	
56	0030.0680 Neg Flow Verify: dP2 at -10,0 (Setpoint 10) [14	SLPM	-14,5 to -5,5	-10,2	√	
57	0030.0690 Neg Flow Verify: dP2 at -5,0 (Setpoint 5) [14m	SLPM	-8,8 to -1,2	-5,1	✓	
58	0030.0700 Neg Flow Verify: dP2 at -0,0 (Setpoint 0) [14m	SLPM	-2,0 to 2,0	-0,0	✓	
59	0030.0950 Contr.Press.Ver.(8302): P1 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,09	√	
60	0030.0960 Monit.Press.Ver.(4232): P2 at 0,00 (SP 0, 1, 1)	cmH20	-1,00 to 1,00	0,00	√	
61	0030.0160 Pprox Press. Ver.(8101): P2 at 0,00 (SP 0, 1, 1	cmH2O	-1,00 to 1,00	0,00	✓	
62	0030.0970 Contr.Press.Ver.(10672): P1 at 3,95 (SP 4, 1, 1	cmH20	2,95 to 4,95	3,99	✓	
63	0030.0980 Monit.Press.Ver.(6990): P2 at 3,95 (SP 4, 1, 1)	cmH20	2,95 to 4,95	3,90	✓	
64	0030.0161 Pprox Press. Ver.(10483): P2 at 3,95 (SP 4, 1,	cmH2O	2,95 to 4,95	3,92	✓	
65	0030.0990 Contr.Press.Ver.(19179): P1 at 17,93 (SP 18, 1,	cmH2O	16,93 to 18,93	18,00	√	
66	0030.1000 Monit.Press.Ver.(16770): P2 at 17,93 (SP 18, 1,	cmH20	16,93 to 18,93	17,85	✓	
67	0030.0162 Pprox Press. Ver.(18969): P2 at 17,93 (SP 18, 1	cmH20	16,93 to 18,93	17,89	√	
68	0030.1010 Contr.Press.Ver.(27682): P1 at 31,92 (SP 32, 1,	cmH20	30,92 to 32,92	31,99	√	
69	0030.1020 Monit.Press.Ver.(26576): P2 at 31,92 (SP 32, 1,	cmH20	30,92 to 32,92	31,82	✓	
70	0030.0163 Pprox Press. Ver.(27472): P2 at 31,92 (SP 32, 1	cmH2O	30,92 to 32,92	31,88	✓	

Test Started On 06/06/24 11:41:03

Elapsed Test Time 48m 5s

Serial Number TV11909271C

Status PASS

Page 2 **of** 6