Test Software P/N 4102483

**Tested Part Number** 1040000

Test Software Name Trilogy FSA (REPAIR TEST)



**Software Revision** 24.0.0.0

Prepared By P. Pascal

## Test Report

Ston	Description	Ilnita	Limita	Dogulto	Pass   Fail
Step	•	Units	Limits	Results	Pass Fail
36	0021.0570 02 Positive Flow Cal: dP1 at 64,7 (SP 65 @82,3P			1666	✓
37	0021.0580 02 Positive Flow Cal: dP1 at 74,3 (SP 75 @81,6P	ADC Cnts	LE 1924	1801	✓
38	0021.0590 O2 Positive Flow Cal: dP1 at 84,2 (SP 85 @80,9P	ADC Cnts	LE 2046	1925	✓
39	0021.0600 02 Positive Flow Cal: dP1 at 94,2 (SP 95 @80,1P	ADC Cnts	LE 2178	2048	✓
40	0021.0610 02 Positive Flow Cal: dP1 at 105,2 (SP 105 @79,	ADC Cnts	LE 2275	2179	✓
41	0021.0620 02 Positive Flow Cal: dP1 at 114,0 (SP 115 @78,	ADC Cnts	LE 2386	2276	✓
42	0021.0630 02 Positive Flow Cal: dP1 at 125,0 (SP 125 @77,	ADC Cnts	LE 2493	2387	✓
43	0021.0640 O2 Positive Flow Cal: dP1 at 135,5 (SP 135 @76,	ADC Cnts	LE 2577	2494	✓
44	0021.0650 02 Positive Flow Cal: dP1 at 144,7 (SP 145 @75,	ADC Cnts	LE 2839	2578	✓
45	0021.0660 O2 Positive Flow Cal: dP1 at 173,7 (SP 175 @72,	\DC Cnts	LE 2972	2840	✓
46	0021.0670 O2 Positive Flow Cal: dP1 at 189,5 (SP 190 @70,	ADC Cnts	LE 3999	2973	✓
47	0021.0880 02 Sensor Cal (NO_O2_ERROR): 539 mV at ambient	N/A	Pass	TRUE	<b>√</b>
48	0021.0930 Write OBM Sensors Cal Table (V: TRUE, A: TRUE)	N/A	Pass	TRUE	<b>√</b>

**Test Started On** 06/10/24 03:30:59

Serial Number TV019111109 Elapsed Test Time 16m 41s

Status PASS

Page 2 **of** 2