Test Software P/N 4102483

Tested Part Number 1054096



Software Revision 24.0.0.0

Prepared By P. Pascal

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0020.0570 Pos Flow Cal: dP2 at 77,0 (Setpoint 75) [15m 33	ADC Cnts	LE 47231	46276	<u> </u>	_
37	0020.0580 Pos Flow Cal: dP2 at 87,0 (Setpoint 85) [15m 33	ADC Cnts	LE 48098	47232	· ✓	
38	0020.0590 Pos Flow Cal: dP2 at 96,8 (Setpoint 95) [15m 33	ADC Cnts	LE 49026	48099	1	
39	0020.0600 Pos Flow Cal: dP2 at 107,2 (Setpoint 105) [15m	ADC Cnts	LE 49783	49027	1	
40	0020.0610 Pos Flow Cal: dP2 at 116,8 (Setpoint 115) [15m	ADC Cnts	LE 50547	49784	√	
41	0020.0620 Pos Flow Cal: dP2 at 127,7 (Setpoint 125) [15m	ADC Cnts	LE 51166	50548	1	
42	0020.0630 Pos Flow Cal: dP2 at 137,0 (Setpoint 135) [15m	ADC Cnts	LE 51886	51167	√	
43	0020.0640 Pos Flow Cal: dP2 at 147,3 (Setpoint 145) [15m	ADC Cnts	LE 53245	51887	√	
44	0020.0650 Pos Flow Cal: dP2 at 168,8 (Setpoint 170) [15m	ADC Cnts	LE 54463	53246	√	
45	0020.0660 Pos Flow Cal: dP2 at 188,7 (Setpoint 190) [15m	ADC Cnts	LE 65435	54464	√	
46	0020.0900 Atmospheric Press Sensors Cal Patm: Slope (m) [lnHg/ADC	EQ 0,000517	0,000517	√	
47	0020.0910 Atmospheric Press Sensors Cal Patm: Intercept (inHg	1,91 to 4,09	2,72	√	
48	0020.0920 Write Sensors Cal Table [16m 41s]	N/A	Pass	TRUE	√	
49	0020.0970 Write Pprox Cal Table [16m 57s]	N/A	Pass	TRUE	√	
50	0020.0935 02 Sensor Cal (not attempted for SW ver.14.2.05	N/A	None	None	√	
51	0020.1030 Write Image Table [17m 18s]	N/A	Pass	TRUE	√	

Test Started On 02/08/24 11:28:11

Serial Number TV118052215 Elapsed Test Time 17m 18s

Status PASS

Page 2 **of** 2