

Test Software P/N 4102483Tested Part Number LA1054096Test Software Name Trilogy FSA (PM TEST)Software Revision 24.0.0.0Prepared By P. Pascal

# Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0030.0280 Pos Flow Verify: dP2 at 47,0 (Setpoint 50) [9m	SLPM	41,6 to 52,4	48,3	✓	
37	0030.0290 Pos Flow Verify: dP2 at 38,0 (Setpoint 40) [9m	SLPM	33,0 to 43,0	38,8	✓	
38	0030.0300 Pos Flow Verify: dP2 at 28,0 (Setpoint 30) [9m	SLPM	23,4 to 32,6	29,0	✓	
39	0030.0310 Pos Flow Verify: dP2 at 20,0 (Setpoint 20) [9m	SLPM	15,0 to 25,0	21,0	✓	
40	0030.0320 Pos Flow Verify: dP2 at 10,0 (Setpoint 10) [9m	SLPM	5,5 to 14,5	11,1	✓	
41	0030.0330 Pos Flow Verify: dP2 at 5,0 (Setpoint 5) [9m 16	SLPM	1,2 to 8,8	5,3	✓	
42	0030.0510 Zero Flow Verify: dP2 at 0,0 (Setpoint 0) [9m 2	SLPM	-2,0 to 2,0	0,0	✓	
43	0030.0550 Neg Flow Verify: dP2 at -138,3 (Setpoint 140) [	SLPM	-147,4 to -129,3	-139,3	✓	
44	0030.0560 Neg Flow Verify: dP2 at -128,3 (Setpoint 130) [	SLPM	-137,0 to -119,7	-128,2	✓	
45	0030.0570 Neg Flow Verify: dP2 at -118,0 (Setpoint 120) [	SLPM	-126,2 to -109,8	-118,2	✓	
46	0030.0580 Neg Flow Verify: dP2 at -107,8 (Setpoint 110) [	SLPM	-115,6 to -100,0	-108,1	✓	
47	0030.0590 Neg Flow Verify: dP2 at -98,8 (Setpoint 100) [1	SLPM	-106,3 to -91,4	-98,8	✓	
48	0030.0600 Neg Flow Verify: dP2 at -88,0 (Setpoint 90) [13	SLPM	-95,0 to -81,0	-87,4	✓	
49	0030.0610 Neg Flow Verify: dP2 at -77,9 (Setpoint 80) [13	SLPM	-84,5 to -71,2	-77,6	✓	
50	0030.0620 Neg Flow Verify: dP2 at -67,2 (Setpoint 70) [13	SLPM	-73,4 to -61,0	-67,5	✓	
51	0030.0630 Neg Flow Verify: dP2 at -57,7 (Setpoint 60) [13	SLPM	-63,5 to -51,9	-58,0	✓	
52	0030.0640 Neg Flow Verify: dP2 at -48,0 (Setpoint 50) [13	SLPM	-53,4 to -42,6	-47,9	✓	
53	0030.0650 Neg Flow Verify: dP2 at -38,0 (Setpoint 40) [13	SLPM	-43,0 to -33,0	-38,2	✓	
54	0030.0660 Neg Flow Verify: dP2 at -28,0 (Setpoint 30) [13	SLPM	-32,6 to -23,4	-27,9	✓	
55	0030.0670 Neg Flow Verify: dP2 at -19,8 (Setpoint 20) [13	SLPM	-24,8 to -14,8	-20,1	✓	
56	0030.0680 Neg Flow Verify: dP2 at -10,0 (Setpoint 10) [13	SLPM	-14,5 to -5,5	-10,4	✓	
57	0030.0690 Neg Flow Verify: dP2 at -5,0 (Setpoint 5) [13m	SLPM	-8,8 to -1,2	-5,0	✓	
58	0030.0700 Neg Flow Verify: dP2 at -0,0 (Setpoint 0) [13m	SLPM	-2,0 to 2,0	0,0	✓	
59	0030.0950 Contr.Press.Ver.(8779): P1 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,00	✓	
60	0030.0960 Monit.Press.Ver.(2496): P2 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,29	✓	
61	0030.0160 Pprox Press. Ver.(9221): P2 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,00	✓	
62	0030.0970 Contr.Press.Ver.(11223): P1 at 3,95 (SP 4, 1, 1)	cmH2O	2,95 to 4,95	3,99	✓	
63	0030.0980 Monit.Press.Ver.(5046): P2 at 3,95 (SP 4, 1, 1)	cmH2O	2,95 to 4,95	3,95	✓	
64	0030.0161 Pprox Press. Ver.(11675): P2 at 3,95 (SP 4, 1,	cmH2O	2,95 to 4,95	3,95	✓	
65	0030.0990 Contr.Press.Ver.(19654): P1 at 17,99 (SP 18, 1,	cmH2O	16,99 to 18,99	17,99	✓	
66	0030.1000 Monit.Press.Ver.(14793): P2 at 17,99 (SP 18, 1,	cmH2O	16,99 to 18,99	17,97	✓	
67	0030.0162 Pprox Press. Ver.(20126): P2 at 17,99 (SP 18, 1	cmH2O	16,99 to 18,99	17,99	✓	
68	0030.1010 Contr.Press.Ver.(28081): P1 at 31,98 (SP 32, 1,	cmH2O	30,98 to 32,98	31,99	✓	
69	0030.1020 Monit.Press.Ver.(24516): P2 at 31,98 (SP 32, 1,	cmH2O	30,98 to 32,98	31,96	✓	
70	0030.0163 Pprox Press. Ver.(28536): P2 at 31,98 (SP 32, 1	cmH2O	30,98 to 32,98	31,98	✓	

Test Started On 04/03/24 11:46:03Serial Number TV113102307Elapsed Test Time 31m 42sStatus **ERROR**

Page 2 of 5

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Production mode.

File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair\_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38  
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20