

Test Software P/N 4102483Tested Part Number LA1054096Test Software Name Trilogy FSA (PM TEST)Software Revision 24.0.0.0Prepared By P. Pascal

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0030.0280 Pos Flow Verify: dP2 at 47,7 (Setpoint 50) [9m	SLPM	42,3 to 53,1	49,4	✓	
37	0030.0290 Pos Flow Verify: dP2 at 37,0 (Setpoint 40) [9m	SLPM	32,0 to 42,0	38,9	✓	
38	0030.0300 Pos Flow Verify: dP2 at 28,0 (Setpoint 30) [9m	SLPM	23,4 to 32,6	29,2	✓	
39	0030.0310 Pos Flow Verify: dP2 at 20,0 (Setpoint 20) [9m	SLPM	15,0 to 25,0	20,5	✓	
40	0030.0320 Pos Flow Verify: dP2 at 10,0 (Setpoint 10) [9m	SLPM	5,5 to 14,5	10,2	✓	
41	0030.0330 Pos Flow Verify: dP2 at 5,0 (Setpoint 5) [9m 32	SLPM	1,2 to 8,8	5,0	✓	
42	0030.0510 Zero Flow Verify: dP2 at 0,0 (Setpoint 0) [9m 4	SLPM	-2,0 to 2,0	-0,1	✓	
43	0030.0550 Neg Flow Verify: dP2 at -138,8 (Setpoint 140) [SLPM	-147,9 to -129,8	-139,8	✓	
44	0030.0560 Neg Flow Verify: dP2 at -127,8 (Setpoint 130) [SLPM	-136,4 to -119,2	-129,1	✓	
45	0030.0570 Neg Flow Verify: dP2 at -117,7 (Setpoint 120) [SLPM	-125,9 to -109,5	-119,0	✓	
46	0030.0580 Neg Flow Verify: dP2 at -108,2 (Setpoint 110) [SLPM	-116,0 to -100,3	-109,4	✓	
47	0030.0590 Neg Flow Verify: dP2 at -98,8 (Setpoint 100) [1	SLPM	-106,3 to -91,4	-99,6	✓	
48	0030.0600 Neg Flow Verify: dP2 at -88,2 (Setpoint 90) [13	SLPM	-95,2 to -81,1	-88,9	✓	
49	0030.0610 Neg Flow Verify: dP2 at -77,0 (Setpoint 80) [13	SLPM	-83,6 to -70,4	-77,8	✓	
50	0030.0620 Neg Flow Verify: dP2 at -68,0 (Setpoint 70) [13	SLPM	-74,2 to -61,8	-68,7	✓	
51	0030.0630 Neg Flow Verify: dP2 at -57,8 (Setpoint 60) [13	SLPM	-63,6 to -52,0	-58,7	✓	
52	0030.0640 Neg Flow Verify: dP2 at -48,0 (Setpoint 50) [13	SLPM	-53,4 to -42,6	-48,2	✓	
53	0030.0650 Neg Flow Verify: dP2 at -38,0 (Setpoint 40) [13	SLPM	-43,0 to -33,0	-38,3	✓	
54	0030.0660 Neg Flow Verify: dP2 at -28,0 (Setpoint 30) [13	SLPM	-32,6 to -23,4	-28,9	✓	
55	0030.0670 Neg Flow Verify: dP2 at -20,0 (Setpoint 20) [13	SLPM	-25,0 to -15,0	-20,6	✓	
56	0030.0680 Neg Flow Verify: dP2 at -10,0 (Setpoint 10) [13	SLPM	-14,5 to -5,5	-10,6	✓	
57	0030.0690 Neg Flow Verify: dP2 at -5,0 (Setpoint 5) [13m	SLPM	-8,8 to -1,2	-5,2	✓	
58	0030.0700 Neg Flow Verify: dP2 at -0,0 (Setpoint 0) [13m	SLPM	-2,0 to 2,0	-0,1	✓	
59	0030.0950 Contr.Press.Ver.(8134): P1 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,11	✓	
60	0030.0960 Monit.Press.Ver.(3550): P2 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,07	✓	
61	0030.0160 Pprox Press. Ver.(8101): P2 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,09	✓	
62	0030.0970 Contr.Press.Ver.(10476): P1 at 3,83 (SP 4, 1, 1)	cmH2O	2,83 to 4,83	3,99	✓	
63	0030.0980 Monit.Press.Ver.(6192): P2 at 3,83 (SP 4, 1, 1)	cmH2O	2,83 to 4,83	3,87	✓	
64	0030.0161 Pprox Press. Ver.(10394): P2 at 3,83 (SP 4, 1,	cmH2O	2,83 to 4,83	3,90	✓	
65	0030.0990 Contr.Press.Ver.(18909): P1 at 17,89 (SP 18, 1,	cmH2O	16,89 to 18,89	17,99	✓	
66	0030.1000 Monit.Press.Ver.(15934): P2 at 17,89 (SP 18, 1,	cmH2O	16,89 to 18,89	17,85	✓	
67	0030.0162 Pprox Press. Ver.(18841): P2 at 17,89 (SP 18, 1	cmH2O	16,89 to 18,89	17,91	✓	
68	0030.1010 Contr.Press.Ver.(27348): P1 at 31,92 (SP 32, 1,	cmH2O	30,92 to 32,92	31,98	✓	
69	0030.1020 Monit.Press.Ver.(25682): P2 at 31,92 (SP 32, 1,	cmH2O	30,92 to 32,92	31,85	✓	
70	0030.0163 Pprox Press. Ver.(27284): P2 at 31,92 (SP 32, 1	cmH2O	30,92 to 32,92	31,93	✓	

Test Started On 01/05/24 09:13:52Serial Number TV119092626Elapsed Test Time 52m 2sStatus **PASS**

Page 2 of 6

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Production mode.
 File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
 Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 1FF8, 05/05/23 08:20