Test Software P/N 4102483

**Tested Part Number** 1040000



**Software Revision** 24.0.0.0

Prepared By P. Pascal

## Test Report

	Test Repo					
Step	Description	Units	Limits	Results	Pass	Fail
36	0021.0570 02 Positive Flow Cal: dP1 at 64,8 (SP 65 @82,5P	ADC Cnts	LE 1907	1773	<b>√</b>	
37	0021.0580 02 Positive Flow Cal: dP1 at 73,8 (SP 75 @81,8P	ADC Cnts	LE 2054	1908	✓	
38	0021.0590 O2 Positive Flow Cal: dP1 at 84,7 (SP 85 @81,0P	ADC Cnts	LE 2166	2055	<b>√</b>	
39	0021.0600 02 Positive Flow Cal: dP1 at 94,7 (SP 95 @80,3P	ADC Cnts	LE 2289	2167	✓	
40	0021.0610 02 Positive Flow Cal: dP1 at 103,5 (SP 105 @79,	ADC Cnts	LE 2406	2290	<b>√</b>	
41	0021.0620 02 Positive Flow Cal: dP1 at 114,7 (SP 115 @78,	ADC Cnts	LE 2509	2407	✓	
42	0021.0630 O2 Positive Flow Cal: dP1 at 124,7 (SP 125 @77,	ADC Cnts	LE 2605	2510	<b>√</b>	
43	0021.0640 02 Positive Flow Cal: dP1 at 134,1 (SP 135 @76,	ADC Cnts	LE 2701	2606	<b>√</b>	
44	0021.0650 02 Positive Flow Cal: dP1 at 144,5 (SP 145 @76,	ADC Cnts	LE 2967	2702	<b>√</b>	
45	0021.0660 O2 Positive Flow Cal: dP1 at 174,5 (SP 175 @72,	ADC Cnts	LE 3081	2968	<b>√</b>	
46	0021.0670 02 Positive Flow Cal: dP1 at 190,2 (SP 190 @71,	ADC Cnts	LE 3999	3082	<b>✓</b>	
47	0021.0880 02 Sensor Cal (NO_O2_ERROR): 574 mV at ambient	N/A	Pass	TRUE	<b>√</b>	
48	0021.0930 Write OBM Sensors Cal Table (V: TRUE, A: TRUE)	N/A	Pass	TRUE	✓	
					-	

**Test Started On** 03/08/24 11:14:51

Serial Number TV018021907 Elapsed Test Time 14m 46s

Status PASS

Page 2 **of** 2