

# Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0030.0570 Neg Flow Verify: dP2 at -118,3 (Setpoint 120) [16m	SLPM	-126,6 to -110,1	-113,0	✓	
37	0030.0580 Neg Flow Verify: dP2 at -108,0 (Setpoint 110) [16m	SLPM	-115,8 to -100,2	-104,0	✓	
38	0030.0590 Neg Flow Verify: dP2 at -98,5 (Setpoint 100) [16m	SLPM	-105,9 to -91,1	-95,2	✓	
39	0030.0600 Neg Flow Verify: dP2 at -89,0 (Setpoint 90) [16m	SLPM	-96,1 to -81,9	-85,1	✓	
40	0030.0610 Neg Flow Verify: dP2 at -78,0 (Setpoint 80) [16m	SLPM	-84,6 to -71,4	-74,0	✓	
41	0030.0620 Neg Flow Verify: dP2 at -68,0 (Setpoint 70) [16m	SLPM	-74,2 to -61,8	-64,6	✓	
42	0030.0630 Neg Flow Verify: dP2 at -58,0 (Setpoint 60) [16m	SLPM	-63,8 to -52,2	-55,5	✓	
43	0030.0640 Neg Flow Verify: dP2 at -48,3 (Setpoint 50) [16m	SLPM	-53,8 to -42,9	-46,1	✓	
44	0030.0650 Neg Flow Verify: dP2 at -38,0 (Setpoint 40) [16m	SLPM	-43,0 to -33,0	-36,4	✓	
45	0030.0660 Neg Flow Verify: dP2 at -28,0 (Setpoint 30) [16m	SLPM	-32,6 to -23,4	-27,1	✓	
46	0030.0670 Neg Flow Verify: dP2 at -20,0 (Setpoint 20) [16m	SLPM	-25,0 to -15,0	-19,2	✓	
47	0030.0680 Neg Flow Verify: dP2 at -10,0 (Setpoint 10) [16m	SLPM	-14,5 to -5,5	-10,1	✓	
48	0030.0690 Neg Flow Verify: dP2 at -5,0 (Setpoint 5) [16m	SLPM	-8,8 to -1,2	-4,7	✓	
49	0030.0700 Neg Flow Verify: dP2 at -0,0 (Setpoint 0) [16m	SLPM	-2,0 to 2,0	-0,0	✓	
50	0030.0950 Contr.Press.Ver.(8125): P1 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,00	✓	
51	0030.0960 Monit.Press.Ver.(3970): P2 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,01	✓	
52	0030.0160 Pprox Press. Ver.(8398): P2 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,02	✓	
53	0030.0970 Contr.Press.Ver.(10587): P1 at 4,05 (SP 4, 1, 1)	cmH2O	3,05 to 5,05	3,99	✓	
54	0030.0980 Monit.Press.Ver.(6794): P2 at 4,05 (SP 4, 1, 1)	cmH2O	3,05 to 5,05	4,04	✓	
55	0030.0161 Pprox Press. Ver.(10815): P2 at 4,05 (SP 4, 1, 1)	cmH2O	3,05 to 5,05	4,00	✓	
56	0030.0990 Contr.Press.Ver.(19070): P1 at 18,07 (SP 18, 1, 1)	cmH2O	17,07 to 19,07	17,99	✓	
57	0030.1000 Monit.Press.Ver.(16573): P2 at 18,07 (SP 18, 1, 1)	cmH2O	17,07 to 19,07	18,01	✓	
58	0030.0162 Pprox Press. Ver.(19276): P2 at 18,07 (SP 18, 1, 1)	cmH2O	17,07 to 19,07	17,93	✓	
59	0030.1010 Contr.Press.Ver.(27561): P1 at 31,99 (SP 32, 1, 1)	cmH2O	30,99 to 32,99	31,98	✓	
60	0030.1020 Monit.Press.Ver.(26322): P2 at 31,99 (SP 32, 1, 1)	cmH2O	30,99 to 32,99	31,94	✓	
61	0030.0163 Pprox Press. Ver.(27773): P2 at 31,99 (SP 32, 1, 1)	cmH2O	30,99 to 32,99	31,92	✓	
62	0030.1030 Contr.Press.Ver.(36041): P1 at 45,95 (SP 46, 1, 1)	cmH2O	44,95 to 46,95	46,00	✓	
63	0030.1040 Monit.Press.Ver.(36109): P2 at 45,95 (SP 46, 1, 1)	cmH2O	44,95 to 46,95	45,90	✓	
64	0030.0164 Pprox Press. Ver.(36280): P2 at 45,95 (SP 46, 1, 1)	cmH2O	44,95 to 46,95	45,92	✓	
65	0030.1050 Contr.Press.Ver.(38460): P1 at 49,95 (SP 50, 1, 1)	cmH2O	48,95 to 50,95	50,00	✓	
66	0030.1060 Monit.Press.Ver.(38904): P2 at 49,95 (SP 50, 1, 1)	cmH2O	48,95 to 50,95	49,89	✓	
67	0030.0165 Pprox Press. Ver.(38693): P2 at 49,95 (SP 50, 1, 1)	cmH2O	48,95 to 50,95	49,94	✓	
68	0030.1070 Contr.Press.Ver.(45867): P1 at 62,10 (SP 65, 1, 1)	cmH2O	61,10 to 63,10	62,19	✓	
69	0030.1080 Monit.Press.Ver.(47453): P2 at 62,10 (SP 65, 1, 1)	cmH2O	61,10 to 63,10	62,13	✓	
70	0030.0166 Pprox Press. Ver.(46095): P2 at 62,10 (SP 65, 1, 1)	cmH2O	61,10 to 63,10	62,12	✓	

Test Started On 08/16/24 11:00:10Serial Number TV018100946Elapsed Test Time 21m 41sStatus **PASS**

Page 2 of 3

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Troubleshooting: Group Level mode.  
 File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair\_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38  
 Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20