Test Software P/N 4102483

Tested Part Number 1040000

Test Software Name Trilogy FSA (REPAIR TEST)



Software Revision 24.0.0.0

Prepared By P. Pascal

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0021.0560 02 Positive Flow Cal: dP1 at 55,0 (SP 55 @79,6P	ADC Cnts	LE 1774	1630	√	
37	0021.0570 02 Positive Flow Cal: dPl at 65,0 (SP 65 @78,8P	ADC Cnts	LE 1915	1775	✓	
38	0021.0580 O2 Positive Flow Cal: dP1 at 75,0 (SP 75 @78,1P.	ADC Cnts	LE 2047	1916	✓	
39	0021.0590 O2 Positive Flow Cal: dP1 at 84,7 (SP 85 @77,5P	ADC Cnts	LE 2172	2048	✓	
40	0021.0600 02 Positive Flow Cal: dP1 at 94,3 (SP 95 @76,9P	ADC Cnts	LE 2299	2173	✓	
41	0021.0610 02 Positive Flow Cal: dP1 at 104,9 (SP 105 @76,	ADC Cnts	LE 2402	2300	√	
42	0021.0620 O2 Positive Flow Cal: dP1 at 114,3 (SP 115 @75,	ADC Cnts	LE 2512	2403	√	
43	0021.0630 O2 Positive Flow Cal: dP1 at 124,9 (SP 125 @74,	ADC Cnts	LE 2607	2513	√	
44	0021.0640 O2 Positive Flow Cal: dP1 at 134,2 (SP 135 @73,	ADC Cnts	LE 2689	2608	√	
45	0021.0650 O2 Positive Flow Cal: dP1 at 144,2 (SP 145 @73,	ADC Cnts	LE 2950	2690	√	
46	0021.0660 O2 Positive Flow Cal: dP1 at 174,3 (SP 175 @70,	ADC Cnts	LE 3082	2950	√	
47	0021.0670 O2 Positive Flow Cal: dP1 at 190,3 (SP 190 @68,	ADC Cnts	LE 3999	3083	 	
48	0021.0880 02 Sensor Cal (NO_O2_ERROR): 557 mV at ambient	N/A	Pass	TRUE	√	
49	0021.0930 Write OBM Sensors Cal Table [18m 52s]	N/A	Pass	FALSE		×
					\bot	
					\perp	

Test Started On 05/13/24 09:35:49

Serial Number TV018022010 Elapsed Test Time 18m 52s

Status FAIL

Page 2 **of** 2