Test Software P/N 4102483 Tested Part Number LA1054096

Test Software Name Trilogy FSA (PM TEST)

Software Revision 24.0.0.0 Prepared By P. Pascal

PHILIPS RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
141	0040.0700 SD Locked Card Inserted: SD Detected (3) [35m 3	N/A	EQ YES	YES	✓	
142	0040.0710 SD Locked Card Inserted: SD Recognized (3) [35m	N/A	EQ YES	YES	✓	
143	0040.0720 SD Locked Card Inserted: Write Protected Mode (	N/A	EQ YES	YES	✓	
144	0040.0730 SD Card Removed: SD Detected (4) [36m 0s]	N/A	EQ NO	NO	✓	
145	0040.0740 SD Card Removed: SD Recognized (4) [36m 0s]	N/A	EQ NO	NO	✓	
146	0040.0750 SD Card Removed: Write Protected Mode (4) [36m	N/A	EQ YES	YES	✓	
147	0040.0780 SD Unlocked Card Inserted: Write Mode Enable (3	N/A	Pass	TRUE	✓	
148	0040.0790 SD: File Write [36m 39s]	N/A	Pass	TRUE	✓	
149	0040.0800 SD: File Length [36m 41s]	N/A	Pass	TRUE	✓	
150	0040.0810 SD: File Read [36m 42s]	N/A	Pass	TRUE	✓	
151	0040.0820 SD: File Erase [36m 43s]	N/A	Pass	TRUE	✓	
152	0040.0840 Nurse Call OFF (NO: 9999,00KOhm, NC: 0,00Ohm) [	N/A	Pass	TRUE	✓	
153	0040.0850 Nurse Call ON (NO: 0,00Ohm, NC: 9999,00KOhm) [3	N/A	Pass	TRUE	✓	
154	0040.0900 P1 with Auto Null Valve Closed at 30,02cmH2O [3	cmH2O	27,02 to 33,02	29,97	✓	
155	0040.0910 P1 with Auto Null Valve Opened at 62,07cmH2O [3	cmH2O	-3,00 to 3,00	0,08	✓	
156	0040.0950 Cooling Fans: Stirring Fan [38m 11s]	N/A	Pass	TRUE	✓	
157	0040.0960 Cooling Fans: Battery Fan [38m 13s]	N/A	Pass	TRUE	✓	
158	0040.0970 Cooling Fans: Exhaust Fan [38m 26s]	N/A	Pass	TRUE	✓	
159	0040.1300 Current Consumption in Sleep Mode (1) [39m 42s]	A	LE 0,20	0,00	✓	
160	0040.1400 O2 Low Flow @ 20,20PSIG [44m 31s]	SLPM	GE 7	15	✓	
161	0040.1410 DSP CRC: JTAG ID Code [44m 50s]	N/A	EQ 01F4001D	01F4001D	✓	
162	0040.1420 DSP CRC: DSP Program CRC [44m 50s]	N/A	EQ FE0F	FE0F	✓	

Test Started On 07/04/24 12:45:34

Serial Number TV116081143 Elapsed Test Time 44m 50s

Status **FAIL**

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Production mode.
File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20