PHILIPS

RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
176	0050.0180 Device's SD Card Inserted: SD Detected (3) [56m	N/A	EQ YES	YES	✓	
177	0050.0190 Device's SD Card Inserted: SD Recognized (3) [5	N/A	EQ YES	YES	✓	
178	0050.0195 Device's SD Card Inserted: SD Unlocked (3) [56m	N/A	EQ YES	YES	✓	
179	0050.0200 Significant Event Log (129 entries): [58m 45s]	N/A	Pass	TRUE	✓	
180	0050.0210 Engineering Debug Log (142 entries): [1 hr 1m	N/A	Pass	TRUE	✓	
181	0050.0020 Reset Operational Hours (1,4 Hrs) [1 hr 1m 12s]	N/A	Pass	TRUE	✓	

Test Started On 03/28/24 08:27:52

Serial Number TV112010957   **Elapsed Test Time** 1 hr 1m 12s

Status **PASS**