Test Software P/N 4102483

Tested Part Number LA1054096

 $\textbf{Test Software Name} \ \underline{\texttt{Trilogy FSA}} \ (\texttt{PM TEST})$



Software Revision 24.0.0.0

Prepared By P. Pascal

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
71	0030.1030 Contr.Press.Ver.(35735): P1 at 44,88 (SP 46, 1,	cmH2O	43,88 to 45,88	45,99		×
72	0030.1040 Monit.Press.Ver.(34843): P2 at 44,88 (SP 46, 1,	cmH20	43,88 to 45,88	45,01	✓	
73	0030.0164 Pprox Press. Ver.(35915): P2 at 44,88 (SP 46, 1	cmH20	43,88 to 45,88	45,28	✓	
74	0030.1050 Contr.Press.Ver.(38139): P1 at 48,89 (SP 50, 1,	cmH20	47,89 to 49,89	49,99		×
75	0030.1060 Monit.Press.Ver.(37636): P2 at 48,89 (SP 50, 1,	cmH20	47,89 to 49,89	49,02	✓	
76	0030.0165 Pprox Press. Ver.(38387): P2 at 48,89 (SP 50, 1	cmH20	47,89 to 49,89	49,19	✓	
77	0030.1070 Contr.Press.Ver.(46091): P1 at 62,20 (SP 65, 1,	cmH20	61,20 to 63,20	63,21		×
78	0030.1080 Monit.Press.Ver.(46816): P2 at 62,20 (SP 65, 1,	cmH20	61,20 to 63,20	62,32	✓	
79	0030.0166 Pprox Press. Ver.(46291): P2 at 62,20 (SP 65, 1	cmH20	61,20 to 63,20	62,48	✓	
80	0030.1150 Atmospheric Pressure Verify: Patm at 28,35,28,3	inHg	27,75 to 28,95	28,19	✓	
81	0030.1250 Airstream Temperature (Tair) at 21,00 [29m 46s]	C deg	16,00 to 36,00	24,48	✓	
						-
						-
						-
						<u> </u>
			<u> </u>			<u> </u>

Test Started On 08/07/24 11:23:11

3

Serial Number TV113031429 Elapsed Test Time 29m 46s

Status FAIL

Page 3 **of**