Test Software P/N 4102483 **Tested Part Number** 1054096

Test Software Name Trilogy FSA (PM TEST)

Software Revision 24.0.0.0 **Prepared By** P. Pascal

PHILIPS RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
71	0030.1030 Contr.Press.Ver.(36798): P1 at 45,81 (SP 46, 1,	cmH2O	44,81 to 46,81	45,99	✓	
72	0030.1040 Monit.Press.Ver.(34967): P2 at 45,81 (SP 46, 1,	cmH2O	44,81 to 46,81	45,90	✓	
73	0030.0164 Pprox Press. Ver.(36469): P2 at 45,81 (SP 46, 1	cmH2O	44,81 to 46,81	45,93	✓	
74	0030.1050 Contr.Press.Ver.(39195): P1 at 49,79 (SP 50, 1,	cmH2O	48,79 to 50,79	49,98	✓	
75	0030.1060 Monit.Press.Ver.(37739): P2 at 49,79 (SP 50, 1,	cmH2O	48,79 to 50,79	49,89	✓	
76	0030.0165 Pprox Press. Ver.(38838): P2 at 49,79 (SP 50, 1	cmH2O	48,79 to 50,79	49,95	✓	
77	0030.1070 Contr.Press.Ver.(47553): P1 at 63,72 (SP 65, 1,	cmH2O	62,72 to 64,72	63,99	✓	
78	0030.1080 Monit.Press.Ver.(47442): P2 at 63,72 (SP 65, 1,	cmH2O	62,72 to 64,72	63,90	✓	
79	0030.0166 Pprox Press. Ver.(47233): P2 at 63,72 (SP 65, 1	cmH2O	62,72 to 64,72	63,98	✓	
80	0030.1150 Atmospheric Pressure Verify: Patm at 28,05,28,0	inHg	27,45 to 28,65	28,01	✓	
81	0030.1250 Airstream Temperature (Tair) at 27,00 [18m 42s]	C deg	22,00 to 42,00	29,84	✓	

Test Started On 01/10/24 12:41:36

Serial Number TV114120511 **Elapsed Test Time** 18m 42s

Status **FAIL**

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Production mode.
File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20