PHILIPS

RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
176	0050.0180 Device's SD Card Inserted: SD Detected (3) [41m	N/A	EQ YES	YES	✓	
177	0050.0190 Device's SD Card Inserted: SD Recognized (3) [4	N/A	EQ YES	YES	✓	
178	0050.0195 Device's SD Card Inserted: SD Unlocked (3) [41m	N/A	EQ YES	YES	✓	
179	0050.0200 Significant Event Log (124 entries): [42m 51s]	N/A	Pass	TRUE	✓	
180	0050.0210 Engineering Debug Log (94 entries): [44m 52s]	N/A	Pass	TRUE	✓	
181	0050.0020 Reset Operational Hours (8731,9 Hrs) [44m 52s]	N/A	Pass	TRUE	✓	

Test Started On <u>09/04/24 04:43:56</u>

Serial Number <u>TV119042617</u>

Elapsed Test Time <u>44m 52s</u>

Status **PASS**