Test Software P/N 4102483 **Tested Part Number** LA1054096

Test Software Name Trilogy FSA (PM TEST)

Software Revision 24.0.0.0 **Prepared By** P. Pascal

PHILIPS RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
1	0010.0003 Write Charger Limiter Table (Limit=75%) [1m 44s	N/A	Pass	TRUE	✓	
2	0010.0020 Check Leak (1) 121 @ 25,0cmH2O: Control Flow Se	SLPM	-5 to 5	1,05	✓	
3	0010.0021 Check Leak (1) @ 25,0cmH2O: Test Setup [4m 41s]	SLPM	-5 to 5	0,00	✓	
4	0010.0070 Check HW Revision [4m 41s]	N/A	EQ 0	0	✓	
5	0010.0080 Check SW Revision (DSP rev. 13) [4m 41s]	N/A	EQ 14.2.05	14.2.05	✓	
6	0010.0110 Check Clock Settings [4m 41s]	s	LE 300	0	✓	
7	0010.0120 Int.Batt.Cap.@(T=25C,SH=97%,CC=75,CF=0,ME=2) [4	%	15 to 85	86,00		✗
8	0010.0121 Det.Batt.Cap.@(T=24C,SH=96%,CC=7,CF=0,ME=1) (1)	%	15 to 85	22,00	✓	
9	0010.0130 Check Ref. Voltage [4m 41s]	mV	2450 to 2550	2482	✓	
10	0010.0140 Motor Temperature [4m 41s]	C Deg	15 to 82	28	✓	
11	0010.0150 Check CPLD SW Revision [4m 41s]	N/A	EQ 12	12	✓	
12	0010.0160 Boot Monitor SW Revision [4m 41s]	N/A	EQ 4.0	4.0	✓	
13	0010.0170 Int. Batt. S/N [4m 41s]	N/A	000011AA	000011AA	✓	
14	0010.0180 Detach. Batt. S/N [4m 41s]	N/A	00003B7F	00003B7F	✓	

Test Started On 01/10/24 04:14:12

Serial Number TV109120118 **Elapsed Test Time** 4m 41s

Status **FAIL**

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Production mode.
File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20