Test Software P/N 4102483 **Tested Part Number** 1054096

Test Software Name Trilogy FSA (PM TEST)

Software Revision 24.0.0.0 **Prepared By** P. Pascal

PHILIPS RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
71	0030.1030 Contr.Press.Ver.(36574): P1 at 45,82 (SP 46, 1,	cmH2O	44,82 to 46,82	45,99	✓	
72	0030.1040 Monit.Press.Ver.(35680): P2 at 45,82 (SP 46, 1,	cmH2O	44,82 to 46,82	45,96	✓	
73	0030.0164 Pprox Press. Ver.(36888): P2 at 45,82 (SP 46, 1	cmH2O	44,82 to 46,82	46,05	✓	
74	0030.1050 Contr.Press.Ver.(38985): P1 at 49,75 (SP 50, 1,	cmH2O	48,75 to 50,75	49,99	✓	
75	0030.1060 Monit.Press.Ver.(38438): P2 at 49,75 (SP 50, 1,	cmH2O	48,75 to 50,75	49,91	✓	
76	0030.0165 Pprox Press. Ver.(39305): P2 at 49,75 (SP 50, 1	cmH2O	48,75 to 50,75	50,01	✓	
77	0030.1070 Contr.Press.Ver.(47802): P1 at 64,15 (SP 65, 1,	cmH2O	63,15 to 65,15	64,61	✓	
78	0030.1080 Monit.Press.Ver.(48600): P2 at 64,15 (SP 65, 1,	cmH2O	63,15 to 65,15	64,44	✓	
79	0030.0166 Pprox Press. Ver.(48009): P2 at 64,15 (SP 65, 1	cmH2O	63,15 to 65,15	64,43	✓	
80	0030.1150 Atmospheric Pressure Verify: Patm at 28,05,28,0	inHg	27,45 to 28,65	28,07	✓	
81	0030.1250 Airstream Temperature (Tair) at 27,34 [24m 4s]	C deg	22,34 to 42,34	30,96	✓	

Test Started On 01/08/24 03:24:42

Serial Number TV110052651 **Elapsed Test Time** 24m 4s

Status **FAIL**

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Production mode.
File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20