Test Software P/N 4102483

**Tested Part Number** 1040000

Test Software Name Trilogy FSA (REPAIR TEST)



**Software Revision** 24.0.0.0

Prepared By P. Pascal

## **Test Report**

Step	Description	Units	Limits	Results	Pass	Fail
36	0021.0570 02 Positive Flow Cal: dP1 at 65,0 (SP 65 @72,6P	ADC Cnts	LE 2255	2065	<b>√</b>	
37	0021.0580 02 Positive Flow Cal: dP1 at 72,7 (SP 75 @70,8P	ADC Cnts	LE 2512	2256	<b>√</b>	
38	0021.0590 02 Positive Flow Cal: dP1 at 82,6 (SP 85 @68,4P	ADC Cnts	LE 2166	2514		×
39	0021.0600 O2 Positive Flow Cal: dP1 at 94,0 (SP 95 @71,6P	ADC Cnts	LE 2288	2168	<b>√</b>	
40	0021.0610 02 Positive Flow Cal: dP1 at 104,3 (SP 105 @70,	ADC Cnts	LE 2404	2289	<b>√</b>	
41	0021.0620 02 Positive Flow Cal: dP1 at 114,7 (SP 115 @69,	ADC Cnts	LE 2503	2405	<b>√</b>	
42	0021.0630 O2 Positive Flow Cal: dP1 at 124,0 (SP 125 @68,	ADC Cnts	LE 2610	2504	<b>√</b>	
43	0021.0640 02 Positive Flow Cal: dP1 at 134,7 (SP 135 @66,	ADC Cnts	LE 2708	2610	<b>√</b>	
44	0021.0650 O2 Positive Flow Cal: dP1 at 145,5 (SP 145 @65,	ADC Cnts	LE 2964	2709	<b>√</b>	
45	0021.0660 02 Positive Flow Cal: dP1 at 174,7 (SP 175 @62,	ADC Cnts	LE 3103	2965	<b>√</b>	_
46	0021.0670 O2 Positive Flow Cal: dP1 at 189,7 (SP 190 @59,	ADC Cnts	LE 3999	3104	<b>√</b>	_
47	0021.0880 02 Sensor Cal (NO_O2_ERROR): 485 mV at ambient	N/A	Pass	TRUE	<b>√</b>	
48	0021.0930 Write OBM Sensors Cal Table [17m 32s]	N/A	Pass	FALSE		×
						_
						_
					-	
					-	
					-	—
						—

**Test Started On** 06/04/24 08:47:26

Serial Number TV015081808 Elapsed Test Time 17m 32s

Status FAIL

Page 2 **of** 2