

Test Software P/N 4102483Tested Part Number LA1054096Test Software Name Trilogy FSA (PM TEST)Software Revision 24.0.0.0Prepared By P. Pascal

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0030.0280 Pos Flow Verify: dP2 at 48,0 (Setpoint 50) [11m	SLPM	42,6 to 53,4	48,9	✓	
37	0030.0290 Pos Flow Verify: dP2 at 37,3 (Setpoint 40) [11m	SLPM	32,3 to 42,3	38,5	✓	
38	0030.0300 Pos Flow Verify: dP2 at 28,0 (Setpoint 30) [11m	SLPM	23,4 to 32,6	28,0	✓	
39	0030.0310 Pos Flow Verify: dP2 at 20,0 (Setpoint 20) [11m	SLPM	15,0 to 25,0	19,6	✓	
40	0030.0320 Pos Flow Verify: dP2 at 10,0 (Setpoint 10) [11m	SLPM	5,5 to 14,5	10,5	✓	
41	0030.0330 Pos Flow Verify: dP2 at 4,2 (Setpoint 5) [11m 4	SLPM	0,4 to 7,9	5,0	✓	
42	0030.0510 Zero Flow Verify: dP2 at 0,0 (Setpoint 0) [11m	SLPM	-2,0 to 2,0	0,2	✓	
43	0030.0550 Neg Flow Verify: dP2 at -138,5 (Setpoint 140) [SLPM	-147,5 to -129,5	-135,4	✓	
44	0030.0560 Neg Flow Verify: dP2 at -128,0 (Setpoint 130) [SLPM	-136,6 to -119,4	-124,2	✓	
45	0030.0570 Neg Flow Verify: dP2 at -118,5 (Setpoint 120) [SLPM	-126,7 to -110,3	-115,6	✓	
46	0030.0580 Neg Flow Verify: dP2 at -108,8 (Setpoint 110) [SLPM	-116,7 to -101,0	-106,8	✓	
47	0030.0590 Neg Flow Verify: dP2 at -98,0 (Setpoint 100) [1	SLPM	-105,4 to -90,6	-96,6	✓	
48	0030.0600 Neg Flow Verify: dP2 at -88,5 (Setpoint 90) [15	SLPM	-95,5 to -81,5	-86,7	✓	
49	0030.0610 Neg Flow Verify: dP2 at -77,9 (Setpoint 80) [15	SLPM	-84,5 to -71,2	-77,0	✓	
50	0030.0620 Neg Flow Verify: dP2 at -67,8 (Setpoint 70) [15	SLPM	-74,0 to -61,6	-67,9	✓	
51	0030.0630 Neg Flow Verify: dP2 at -57,0 (Setpoint 60) [15	SLPM	-62,8 to -51,2	-58,4	✓	
52	0030.0640 Neg Flow Verify: dP2 at -48,0 (Setpoint 50) [15	SLPM	-53,4 to -42,6	-49,4	✓	
53	0030.0650 Neg Flow Verify: dP2 at -38,0 (Setpoint 40) [15	SLPM	-43,0 to -33,0	-39,5	✓	
54	0030.0660 Neg Flow Verify: dP2 at -28,0 (Setpoint 30) [15	SLPM	-32,6 to -23,4	-29,5	✓	
55	0030.0670 Neg Flow Verify: dP2 at -20,0 (Setpoint 20) [15	SLPM	-25,0 to -15,0	-21,2	✓	
56	0030.0680 Neg Flow Verify: dP2 at -10,0 (Setpoint 10) [15	SLPM	-14,5 to -5,5	-9,9	✓	
57	0030.0690 Neg Flow Verify: dP2 at -5,0 (Setpoint 5) [15m	SLPM	-8,8 to -1,2	-4,5	✓	
58	0030.0700 Neg Flow Verify: dP2 at -0,0 (Setpoint 0) [15m	SLPM	-2,0 to 2,0	0,2	✓	
59	0030.0950 Contr.Press.Ver.(8556): P1 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	1,00	✓	
60	0030.0960 Monit.Press.Ver.(3376): P2 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,12	✓	
61	0030.0160 Pprox Press. Ver.(7797): P2 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,06	✓	
62	0030.0970 Contr.Press.Ver.(10360): P1 at 3,02 (SP 4, 1, 1)	cmH2O	2,02 to 4,02	3,99	✓	
63	0030.0980 Monit.Press.Ver.(5431): P2 at 3,02 (SP 4, 1, 1)	cmH2O	2,02 to 4,02	3,06	✓	
64	0030.0161 Pprox Press. Ver.(9608): P2 at 3,02 (SP 4, 1, 1)	cmH2O	2,02 to 4,02	3,08	✓	
65	0030.0990 Contr.Press.Ver.(18806): P1 at 17,07 (SP 18, 1,	cmH2O	16,07 to 18,07	18,00	✓	
66	0030.1000 Monit.Press.Ver.(15174): P2 at 17,07 (SP 18, 1,	cmH2O	16,07 to 18,07	17,03	✓	
67	0030.0162 Pprox Press. Ver.(18058): P2 at 17,07 (SP 18, 1	cmH2O	16,07 to 18,07	17,06	✓	
68	0030.1010 Contr.Press.Ver.(27262): P1 at 31,08 (SP 32, 1,	cmH2O	30,08 to 32,08	31,99	✓	
69	0030.1020 Monit.Press.Ver.(24927): P2 at 31,08 (SP 32, 1,	cmH2O	30,08 to 32,08	30,95	✓	
70	0030.0163 Pprox Press. Ver.(26492): P2 at 31,08 (SP 32, 1	cmH2O	30,08 to 32,08	30,99	✓	

Test Started On 04/23/24 08:55:05Serial Number TV116022461Elapsed Test Time 54m 18sStatus **PASS**

Page 2 of 6

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Production mode.

File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20