Test Software P/N 4102483 **Tested Part Number** LA1054096

Test Software Name Trilogy FSA (PM TEST)

Software Revision 24.0.0.0 **Prepared By** P. Pascal

PHILIPS RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
71	0030.1030 Contr.Press.Ver.(35844): P1 at 46,02 (SP 46, 1,	cmH2O	45,02 to 47,02	46,00	✓	
72	0030.1040 Monit.Press.Ver.(36302): P2 at 46,02 (SP 46, 1,	cmH2O	45,02 to 47,02	45,83	✓	
73	0030.0164 Pprox Press. Ver.(36318): P2 at 46,02 (SP 46, 1	cmH2O	45,02 to 47,02	45,95	✓	
74	0030.1050 Contr.Press.Ver.(38272): P1 at 50,01 (SP 50, 1,	cmH2O	49,01 to 51,01	49,99	✓	
75	0030.1060 Monit.Press.Ver.(39087): P2 at 50,01 (SP 50, 1,	cmH2O	49,01 to 51,01	49,82	✓	
76	0030.0165 Pprox Press. Ver.(38714): P2 at 50,01 (SP 50, 1	cmH2O	49,01 to 51,01	49,89	✓	
77	0030.1070 Contr.Press.Ver.(45830): P1 at 62,62 (SP 65, 1,	cmH2O	61,62 to 63,62	62,54	✓	
78	0030.1080 Monit.Press.Ver.(47876): P2 at 62,62 (SP 65, 1,	cmH2O	61,62 to 63,62	62,38	✓	
79	0030.0166 Pprox Press. Ver.(46273): P2 at 62,62 (SP 65, 1	cmH2O	61,62 to 63,62	62,42	✓	
80	0030.1150 Atmospheric Pressure Verify: Patm at 28,05,28,0	inHg	27,45 to 28,65	28,36	✓	
81	0030.1250 Airstream Temperature (Tair) at 26,00 [24m 40s]	C deg	21,00 to 41,00	30,55	✓	

Test Started On 01/16/24 09:55:56

Serial Number TV117060619 **Elapsed Test Time** 24m 40s

Status **FAIL**

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Production mode.
File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20