Test Software P/N 4102483

Tested Part Number 1040000

Test Software Name Trilogy FSA (REPAIR TEST)



Software Revision 24.0.0.0

Prepared By P. Pascal

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0021.0570 02 Positive Flow Cal: dPl at 64,0 (SP 65 @76,6P		LE 4	1 4		
37	0021.0580 O2 Positive Flow Cal: dFl at 74,0 (SP 75 @75,0P		LE 4	4	√	
38	0021.0590 O2 Positive Flow Cal: dFl at 74,0 (SF 75 @73,0F		LE 3	4	√	
39	0021.0600 O2 Positive Flow Cal: dFl at 94,0 (SF 85 @73,3F		LE 4	4		<u>*</u>
40	0021.0610 O2 Positive Flow Cal: dF1 at 94,3 (SF 95 @72,3F		LE 3	5	✓	
	0021.0620 02 Positive Flow Cal: dPl at 104,2 (SP 105 @70,					<u>*</u>
41			LE 3	4		<u>×</u>
42	0021.0630 02 Positive Flow Cal: dP1 at 125,0 (SP 125 @69,		LE 3	4		<u>×</u>
43	0021.0640 02 Positive Flow Cal: dP1 at 134,7 (SP 135 @67,		LE 4	4	✓	
44	0021.0650 O2 Positive Flow Cal: dP1 at 144,8 (SP 145 @67,		LE 4	4	✓	
45	0021.0660 O2 Positive Flow Cal: dP1 at 174,0 (SP 175 @63,		LE 4	4	✓	
46	0021.0670 O2 Positive Flow Cal: dP1 at 191,0 (SP 190 @60,		LE 3999	5	✓	
47	0021.0880 02 Sensor Cal (NO_O2_ERROR): 749 mV at ambient	N/A	Pass	TRUE	✓	
48	0021.0930 Write OBM Sensors Cal Table [13m 55s]	N/A	Pass	FALSE		×
						_
						_
						_
						-
						-
						—

Test Started On 02/29/24 11:30:04

Serial Number TV015081808 Elapsed Test Time 13m 55s

Status FAIL

Page 2 **of** 2