

# Test Report

Step	Description	Units	Limits	Results	Pass	Fail
1	0020.0001 Ambient Air: RH=43,40%; T=23,40Cdeg [31s]	N/A	Pass	TRUE	✓	
2	0020.0010 O2 Sensor Heater On (not attempted for SW ver.1.1)	N/A	None	None	✓	
3	0020.0010 Press Sensors Cal P1@4,02&50,01cmH2O/9719&34029	20/ADC C	0,00151 to 0,0018	0,00167	✓	
4	0020.0020 Press Sensors Cal P1@4,02&50,01cmH2O/9719&34029	cmH2O	-18,22 to -9,05	-13,27	✓	
5	0020.0030 Press Sensors Cal P1@4,02&50,01cmH2O/9719&34029	ADC Cnts	5994 to 9998	7988	✓	
6	0020.0040 Press Sensors Cal P2@4,02&50,01cmH2O/9719&34029	20/ADC C	0,00114 to 0,0018	0,00144	✓	
7	0020.0050 Press Sensors Cal P2@4,02&50,01cmH2O/9719&34029	cmH2O	-19,53 to 0,00	-6,02	✓	
8	0020.0060 Press Sensors Cal P2@4,02&50,01cmH2O/9719&34029	ADC Cnts	0 to 10399	4177	✓	
9	0020.0950 Press Sensors Cal Pprox@4,02&50,01cmH2O/9719&34	20/ADC C	0,00151 to 0,0018	0,00166	✓	
10	0020.0960 Press Sensors Cal Pprox@4,02&50,01cmH2O/9719&34	cmH2O	-18,22 to -9,05	-13,41	✓	
11	0020.0970 Press Sensors Cal Pprox@4,02&50,01cmH2O/9719&34	ADC Cnts	5994 to 9998	8093	✓	
12	0020.0120 Neg Flow Cal: dP2 at -146,3 (Setpoint 145) [8m	ADC Cnts	GE 101	13827	✓	
13	0020.0130 Neg Flow Cal: dP2 at -132,8 (Setpoint 135) [8m	ADC Cnts	GE 13828	14746	✓	
14	0020.0140 Neg Flow Cal: dP2 at -123,3 (Setpoint 125) [8m	ADC Cnts	GE 14746	15440	✓	
15	0020.0150 Neg Flow Cal: dP2 at -113,5 (Setpoint 115) [8m	ADC Cnts	GE 15441	16233	✓	
16	0020.0160 Neg Flow Cal: dP2 at -103,8 (Setpoint 105) [8m	ADC Cnts	GE 16234	17069	✓	
17	0020.0170 Neg Flow Cal: dP2 at -93,7 (Setpoint 95) [8m 52	ADC Cnts	GE 17070	18087	✓	
18	0020.0180 Neg Flow Cal: dP2 at -83,0 (Setpoint 85) [8m 52	ADC Cnts	GE 18088	19189	✓	
19	0020.0190 Neg Flow Cal: dP2 at -72,8 (Setpoint 75) [8m 52	ADC Cnts	GE 19190	20279	✓	
20	0020.0200 Neg Flow Cal: dP2 at -63,3 (Setpoint 65) [8m 52	ADC Cnts	GE 20280	21364	✓	
21	0020.0210 Neg Flow Cal: dP2 at -53,0 (Setpoint 55) [8m 52	ADC Cnts	GE 21365	22642	✓	
22	0020.0220 Neg Flow Cal: dP2 at -43,0 (Setpoint 45) [8m 52	ADC Cnts	GE 22644	24028	✓	
23	0020.0230 Neg Flow Cal: dP2 at -33,0 (Setpoint 35) [8m 52	ADC Cnts	GE 24029	25482	✓	
24	0020.0240 Neg Flow Cal: dP2 at -23,0 (Setpoint 25) [8m 52	ADC Cnts	GE 25483	27124	✓	
25	0020.0250 Neg Flow Cal: dP2 at -14,8 (Setpoint 15) [8m 52	ADC Cnts	GE 27125	28769	✓	
26	0020.0260 Neg Flow Cal: dP2 at -5,0 (Setpoint 5) [8m 52s]	ADC Cnts	GE 28770	31720	✓	
27	0020.0270 Neg Flow Cal: dP2 at -0,0 (Setpoint 0) [8m 52s]	ADC Cnts	GE 31721	33145	✓	
28	0020.0450 Raw Zero Flow: dP2 at 0,0 (Setpoint 0) [8m 52s]	ADC Cnts	GE 0	33145	✓	
29	0020.0500 Pos Flow Cal: dP2 at 5,0 (Setpoint 5) [14m 19s]	ADC Cnts	LE 37821	34706	✓	
30	0020.0510 Pos Flow Cal: dP2 at 15,0 (Setpoint 15) [14m 19	ADC Cnts	LE 39632	37822	✓	
31	0020.0520 Pos Flow Cal: dP2 at 24,0 (Setpoint 25) [14m 19	ADC Cnts	LE 41759	39633	✓	
32	0020.0530 Pos Flow Cal: dP2 at 38,0 (Setpoint 35) [14m 19	ADC Cnts	LE 42899	41760	✓	
33	0020.0540 Pos Flow Cal: dP2 at 47,0 (Setpoint 45) [14m 19	ADC Cnts	LE 44166	42900	✓	
34	0020.0550 Pos Flow Cal: dP2 at 57,0 (Setpoint 55) [14m 19	ADC Cnts	LE 45302	44167	✓	
35	0020.0560 Pos Flow Cal: dP2 at 67,0 (Setpoint 65) [14m 19	ADC Cnts	LE 46367	45303	✓	

Test Started On 03/25/24 08:41:07Serial Number TV118062688Elapsed Test Time 15m 20sStatus **PASS**

Page 1 of 2

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Troubleshooting: Group Level mode.  
 File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair\_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38  
 Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 77D6, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20