

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0030.0280 Pos Flow Verify: dP2 at 47,2 (Setpoint 50) [9m	SLPM	41,8 to 52,6	46,7	✓	
37	0030.0290 Pos Flow Verify: dP2 at 37,7 (Setpoint 40) [9m	SLPM	32,7 to 42,7	37,2	✓	
38	0030.0300 Pos Flow Verify: dP2 at 27,3 (Setpoint 30) [9m	SLPM	22,7 to 31,9	28,0	✓	
39	0030.0310 Pos Flow Verify: dP2 at 19,5 (Setpoint 20) [9m	SLPM	14,5 to 24,5	20,1	✓	
40	0030.0320 Pos Flow Verify: dP2 at 10,0 (Setpoint 10) [9m	SLPM	5,5 to 14,5	10,6	✓	
41	0030.0330 Pos Flow Verify: dP2 at 5,0 (Setpoint 5) [9m 29	SLPM	1,2 to 8,8	4,9	✓	
42	0030.0510 Zero Flow Verify: dP2 at 0,0 (Setpoint 0) [9m 3	SLPM	-2,0 to 2,0	0,0	✓	
43	0030.0550 Neg Flow Verify: dP2 at -138,8 (Setpoint 140) [SLPM	-147,9 to -129,8	-135,7	✓	
44	0030.0560 Neg Flow Verify: dP2 at -127,5 (Setpoint 130) [SLPM	-136,1 to -118,9	-125,2	✓	
45	0030.0570 Neg Flow Verify: dP2 at -118,3 (Setpoint 120) [SLPM	-126,6 to -110,1	-115,8	✓	
46	0030.0580 Neg Flow Verify: dP2 at -108,5 (Setpoint 110) [SLPM	-116,3 to -100,7	-106,3	✓	
47	0030.0590 Neg Flow Verify: dP2 at -98,0 (Setpoint 100) [1	SLPM	-105,4 to -90,6	-96,4	✓	
48	0030.0600 Neg Flow Verify: dP2 at -88,9 (Setpoint 90) [13	SLPM	-95,9 to -81,8	-86,6	✓	
49	0030.0610 Neg Flow Verify: dP2 at -77,5 (Setpoint 80) [13	SLPM	-84,1 to -70,9	-75,9	✓	
50	0030.0620 Neg Flow Verify: dP2 at -67,7 (Setpoint 70) [13	SLPM	-73,9 to -61,5	-66,6	✓	
51	0030.0630 Neg Flow Verify: dP2 at -57,0 (Setpoint 60) [13	SLPM	-62,8 to -51,2	-56,7	✓	
52	0030.0640 Neg Flow Verify: dP2 at -48,3 (Setpoint 50) [13	SLPM	-53,8 to -42,9	-47,2	✓	
53	0030.0650 Neg Flow Verify: dP2 at -38,0 (Setpoint 40) [13	SLPM	-43,0 to -33,0	-37,1	✓	
54	0030.0660 Neg Flow Verify: dP2 at -28,0 (Setpoint 30) [13	SLPM	-32,6 to -23,4	-27,6	✓	
55	0030.0670 Neg Flow Verify: dP2 at -19,2 (Setpoint 20) [13	SLPM	-24,2 to -14,2	-19,2	✓	
56	0030.0680 Neg Flow Verify: dP2 at -10,0 (Setpoint 10) [13	SLPM	-14,5 to -5,5	-9,8	✓	
57	0030.0690 Neg Flow Verify: dP2 at -5,0 (Setpoint 5) [13m	SLPM	-8,8 to -1,2	-4,9	✓	
58	0030.0700 Neg Flow Verify: dP2 at -0,0 (Setpoint 0) [13m	SLPM	-2,0 to 2,0	0,0	✓	
59	0030.0950 Contr.Press.Ver.(8004): P1 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,06	✓	
60	0030.0960 Monit.Press.Ver.(4312): P2 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,00	✓	
61	0030.0160 Pprox Press. Ver.(8137): P2 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,02	✓	
62	0030.0970 Contr.Press.Ver.(10368): P1 at 3,93 (SP 4, 1, 1)	cmH2O	2,93 to 4,93	3,99	✓	
63	0030.0980 Monit.Press.Ver.(7036): P2 at 3,93 (SP 4, 1, 1)	cmH2O	2,93 to 4,93	3,91	✓	
64	0030.0161 Pprox Press. Ver.(10493): P2 at 3,93 (SP 4, 1,	cmH2O	2,93 to 4,93	3,95	✓	
65	0030.0990 Contr.Press.Ver.(18774): P1 at 17,91 (SP 18, 1,	cmH2O	16,91 to 18,91	17,99	✓	
66	0030.1000 Monit.Press.Ver.(16718): P2 at 17,91 (SP 18, 1,	cmH2O	16,91 to 18,91	17,87	✓	
67	0030.0162 Pprox Press. Ver.(18859): P2 at 17,91 (SP 18, 1	cmH2O	16,91 to 18,91	17,87	✓	
68	0030.1010 Contr.Press.Ver.(27180): P1 at 31,90 (SP 32, 1,	cmH2O	30,90 to 32,90	31,99	✓	
69	0030.1020 Monit.Press.Ver.(26429): P2 at 31,90 (SP 32, 1,	cmH2O	30,90 to 32,90	31,87	✓	
70	0030.0163 Pprox Press. Ver.(27290): P2 at 31,90 (SP 32, 1	cmH2O	30,90 to 32,90	31,89	✓	

Test Started On 03/25/24 11:46:39Serial Number TV118062688Elapsed Test Time 1 hr 55m 32sStatus **ERROR**

Page 2 of 6

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Production mode.
 File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
 Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20