Test Software P/N 4102483 Te

Tested Part Number 1040000

 $\textbf{Test Software Name} \ \underline{\texttt{Trilogy FSA}} \ (\texttt{PM TEST})$



Software Revision 24.0.0.0

Prepared By P. Pascal

Test Report

	Test Hepott						
Step	Description	Units	Limits	Results	Pass	Fail	
71	0030.1150 Atmospheric Pressure Verify: Patm at 28,05,28,0	inHg	27,45 to 28,65	28,01	√		
72	0030.1250 Airstream Temperature (Tair) at 28,00 [32m 50s]	C deg	23,00 to 43,00	33,30	✓		
						_	
						_	
						_	
						_	
						_	
			<u> </u>		Ш		

Test Started On 02/02/24 09:14:05

Serial Number TV018100939 Elapsed Test Time 32m 50s

Status PASS

Page 3 **of** 3