Test Software P/N 4102483

Tested Part Number LA1054096

Test Software Name Trilogy FSA (PM TEST)



Software Revision 24.0.0.0

Prepared By P. Pascal

Test Report

	<u> </u>	_ •				
Step	Description	Units	Limits	Results	Pass	Fail
1	0040.0610 Act Exh Purge ValveOpened (Dump) at 29,91cmH20	cmH20	-1.0 to 1.0	0,04	√	
2	0040.0600 Act Exh Purge ValveClosed at 30,00cmH20 (1) [2m	cmH20	27,00 to 33,00	29,87	✓	
3	0040.0620 Act Exh Proport Valve Opened 0% at 29,96cmH20 [cmH20	-1.0 to 10.0	0,09	√	
4	0040.0630 Act Exh Proport Valve Opened 100% at 30,00cmH20	cmH20	27,0 to 33,0	29,85	✓	
5	0040.0640 Act Exh Proport Valve Opened 0% at 30,02cmH20 [cmH20	-1.0 to 10.0	0,08	✓	
						L
						<u> </u>
						L

Test Started On 03/25/24 11:13:41

Serial Number TV118062688 Elapsed Test Time 3m 3s

Status PASS

Page 1 **of** 1