Test Software P/N 4102483

Tested Part Number LA1054096

Test Software Name Trilogy FSA (PM TEST)



Software Revision 24.0.0.0

Prepared By P. Pascal

Test Report

Gu	Provided	l #74	I I	D14.	ln.	177. *1
Step	Description	Units	Limits	Results	Pass	Fail
1	0040.0610 Act Exh Purge ValveOpened (Dump) at 29,92cmH20	cmH20	-1.0 to 1.0	0,05	✓	
2	0040.0600 Act Exh Purge ValveClosed at 30,00cmH2O (2) [3m	cmH20	27,00 to 33,00	22,37		×
3	0040.0620 Act Exh Proport Valve Opened 0% at 29,99cmH20 [cmH20	-1.0 to 10.0	0,09	✓	
4	0040.0630 Act Exh Proport Valve Opened 100% at 30,01cmH20	cmH2O	27,0 to 33,0	22,36		×
5	0040.0640 Act Exh Proport Valve Opened 0% at 29,98cmH20 [cmH20	-1.0 to 10.0	0,08	✓	
						-
						_
					-	
					_	
						_
					_	
						L
		-	+		-	—

Test Started On 03/25/24 10:07:37

Serial Number TV118062688 Elapsed Test Time 4m 6s

Status FAIL

Page 1 **of** 1