Test Software P/N 4102483 **Tested Part Number** 1054096

Test Software Name Trilogy FSA (PM TEST)

Software Revision 24.0.0.0 **Prepared By** P. Pascal

PHILIPS RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
71	0030.1030 Contr.Press.Ver.(36616): P1 at 45,81 (SP 46, 1,	cmH2O	44,81 to 46,81	46,00	✓	
72	0030.1040 Monit.Press.Ver.(35738): P2 at 45,81 (SP 46, 1,	cmH2O	44,81 to 46,81	46,01	✓	
73	0030.0164 Pprox Press. Ver.(36964): P2 at 45,81 (SP 46, 1	cmH2O	44,81 to 46,81	45,99	✓	
74	0030.1050 Contr.Press.Ver.(39026): P1 at 49,75 (SP 50, 1,	cmH2O	48,75 to 50,75	49,99	✓	
75	0030.1060 Monit.Press.Ver.(38517): P2 at 49,75 (SP 50, 1,	cmH2O	48,75 to 50,75	49,96	✓	
76	0030.0165 Pprox Press. Ver.(39325): P2 at 49,75 (SP 50, 1	cmH2O	48,75 to 50,75	49,97	✓	
77	0030.1070 Contr.Press.Ver.(47385): P1 at 63,50 (SP 65, 1,	cmH2O	62,50 to 64,50	63,90	✓	
78	0030.1080 Monit.Press.Ver.(48201): P2 at 63,50 (SP 65, 1,	cmH2O	62,50 to 64,50	63,74	✓	
79	0030.0166 Pprox Press. Ver.(47642): P2 at 63,50 (SP 65, 1	cmH2O	62,50 to 64,50	63,62	✓	
80	[20m 28s]			TERMINATED		✘
81	0030.1150 Atmospheric Pressure Verify: Patm at 28,05,28,0	inHg	27,45 to 28,65	28,07	✓	
82	[21m 46s]			TERMINATED		✘
83	0030.1250 Airstream Temperature (Tair) at 30,00 [22m 23s]	C deg	25,00 to 45,00	34,20	✓	

Test Started On 01/08/24 04:13:53

Serial Number TV110052651 **Elapsed Test Time** 22m 23s

Status **TERMINATED**

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Production mode.
File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20