Test Software P/N 4102483

Tested Part Number 1040000

Test Software Name Trilogy FSA (REPAIR TEST)



Software Revision 24.0.0.0

Prepared By P. Pascal

Test Report

Ctan I	Description	l IInita	I imita	Daggal4a	Decal East
Step	Description	Units	Limits	Results	Pass Fail
36	0021.0570 O2 Positive Flow Cal: dP1 at 64,3 (SP 65 @82,3P	ADC Cnts	LE 1924	1788	✓
37	0021.0580 O2 Positive Flow Cal: dP1 at 74,2 (SP 75 @82,1P	ADC Cnts	LE 2060	1925	✓
38	0021.0590 O2 Positive Flow Cal: dP1 at 84,3 (SP 85 @81,5P	ADC Cnts	LE 2178	2061	✓
39	0021.0600 O2 Positive Flow Cal: dP1 at 94,2 (SP 95 @80,9P	ADC Cnts	LE 2315	2178	✓
40	0021.0610 02 Positive Flow Cal: dP1 at 105,0 (SP 105 @80,	ADC Cnts	LE 2429	2316	✓
41	0021.0620 O2 Positive Flow Cal: dP1 at 114,8 (SP 115 @79,	ADC Cnts	LE 2533	2430	✓
42	0021.0630 O2 Positive Flow Cal: dP1 at 124,7 (SP 125 @78,	ADC Cnts	LE 2634	2534	√
43	0021.0640 O2 Positive Flow Cal: dP1 at 134,3 (SP 135 @77,	ADC Cnts	LE 2728	2635	✓
44	0021.0650 02 Positive Flow Cal: dP1 at 145,0 (SP 145 @76,	ADC Cnts	LE 2988	2729	✓
45	0021.0660 O2 Positive Flow Cal: dP1 at 175,3 (SP 175 @73,	\DC Cnts	LE 3121	2989	√
46	0021.0670 02 Positive Flow Cal: dP1 at 190,0 (SP 190 @71,	ADC Cnts	LE 3999	3122	√
47	0021.0880 02 Sensor Cal (NO_O2_ERROR): 708 mV at ambient	N/A	Pass	TRUE	√
48	0021.0930 Write OBM Sensors Cal Table (V: TRUE, A: TRUE)	N/A	Pass	TRUE	√
					+
					+
					- - -
					- - -
					- - -
					- - -
					- - -
					+
					+
					-

Test Started On 06/07/24 12:03:58

Serial Number TV012071003

Elapsed Test Time 16m 25s

Status PASS

Page 2 **of** 2