Test Software P/N 4102483 Tested Part Number LA1054096

Test Software Name Trilogy FSA (PM TEST)

Software Revision 24.0.0.0 Prepared By P. Pascal

PHILIPS RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
71	0030.1030 Contr.Press.Ver.(35851): P1 at 46,05 (SP 46, 1,	cmH2O	45,05 to 47,05	45,99	✓	
72	0030.1040 Monit.Press.Ver.(36303): P2 at 46,05 (SP 46, 1,	cmH2O	45,05 to 47,05	45,82	✓	
73	0030.0164 Pprox Press. Ver.(36327): P2 at 46,05 (SP 46, 1	cmH2O	45,05 to 47,05	45,88	✓	
74	0030.1050 Contr.Press.Ver.(38273): P1 at 50,04 (SP 50, 1,	cmH2O	49,04 to 51,04	49,98	✓	
75	0030.1060 Monit.Press.Ver.(39090): P2 at 50,04 (SP 50, 1,	cmH2O	49,04 to 51,04	49,80	✓	
76	0030.0165 Pprox Press. Ver.(38738): P2 at 50,04 (SP 50, 1	cmH2O	49,04 to 51,04	49,88	✓	
77	0030.1070 Contr.Press.Ver.(45864): P1 at 62,64 (SP 65, 1,	cmH2O	61,64 to 63,64	62,54	✓	
78	0030.1080 Monit.Press.Ver.(47868): P2 at 62,64 (SP 65, 1,	cmH2O	61,64 to 63,64	62,35	✓	
79	0030.0166 Pprox Press. Ver.(46301): P2 at 62,64 (SP 65, 1	cmH2O	61,64 to 63,64	62,47	✓	
80	0030.1150 Atmospheric Pressure Verify: Patm at 28,05,28,0	inHg	27,45 to 28,65	28,37	✓	
81	0030.1250 Airstream Temperature (Tair) at 26,00 [19m 33s]	C deg	21,00 to 41,00	30,41	✓	

Test Started On 01/16/24 09:34:37

Serial Number TV117060619 Elapsed Test Time 19m 33s

Status **FAIL**

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Production mode.
File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20