

Test Report

| Step | Description | Units | Limits | Results | Pass | Fail |
|------|---|-------|------------------|---------------|------|------|
| 71 | 0030.1030 Contr.Press.Ver.(35872): P1 at 45,84 (SP 46, 1, | cmH2O | 44,84 to 46,84 | 46,00 | ✓ | |
| 72 | 0030.1040 Monit.Press.Ver.(36216): P2 at 45,84 (SP 46, 1, | cmH2O | 44,84 to 46,84 | 46,02 | ✓ | |
| 73 | 0030.0164 Pprox Press. Ver.(35805): P2 at 45,84 (SP 46, 1 | cmH2O | 44,84 to 46,84 | 46,03 | ✓ | |
| 74 | 0030.1050 Contr.Press.Ver.(38288): P1 at 49,85 (SP 50, 1, | cmH2O | 48,85 to 50,85 | 50,00 | ✓ | |
| 75 | 0030.1060 Monit.Press.Ver.(39010): P2 at 49,85 (SP 50, 1, | cmH2O | 48,85 to 50,85 | 50,05 | ✓ | |
| 76 | 0030.0165 Pprox Press. Ver.(38190): P2 at 49,85 (SP 50, 1 | cmH2O | 48,85 to 50,85 | 50,07 | ✓ | |
| 77 | 0030.1070 Contr.Press.Ver.(46372): P1 at 63,35 (SP 65, 1, | cmH2O | 62,35 to 64,35 | 63,39 | ✓ | |
| 78 | 0030.1080 Monit.Press.Ver.(48363): P2 at 63,35 (SP 65, 1, | cmH2O | 62,35 to 64,35 | 63,51 | ✓ | |
| 79 | 0030.0166 Pprox Press. Ver.(46338): P2 at 63,35 (SP 65, 1 | cmH2O | 62,35 to 64,35 | 63,51 | ✓ | |
| 80 | 0030.1150 Atmospheric Pressure Verify: Patm at 28,05,28,0 | inHg | 27,45 to 28,65 | 28,40 | ✓ | |
| 81 | 0030.1250 Airstream Temperature (Tair) at 26,00 [18m 43s] | C deg | 21,00 to 41,00 | 29,21 | ✓ | |
| 82 | 0040.0010 Keypad Backlight Test [18m 47s] | N/A | Pass | TRUE | ✓ | |
| 83 | 0040.0020 LCD Backlight Test [18m 50s] | N/A | Pass | TRUE | ✓ | |
| 84 | 0040.0030 Primary Audible Alarm Soft [18m 56s] | N/A | Pass | TRUE | ✓ | |
| 85 | 0040.0040 Primary Audible Alarm Soft heard by HW (1) [18m | N/A | Pass | TRUE | ✓ | |
| 86 | 0040.0050 Primary Audible Alarm Loud [18m 56s] | N/A | Pass | TRUE | ✓ | |
| 87 | 0040.0060 Primary Audible Alarm Loud heard by HW (1) [18m | N/A | Pass | TRUE | ✓ | |
| 88 | 0040.0070 Red Alarm LED [19m 0s] | N/A | Pass | TRUE | ✓ | |
| 89 | 0040.0080 Yellow Alarm LED [19m 0s] | N/A | Pass | TRUE | ✓ | |
| 90 | 0040.0090 Secondary Audible Alarm Soft [19m 7s] | N/A | Pass | TRUE | ✓ | |
| 91 | 0040.0100 Secondary Audible Alarm Soft heard by HW (1) [1 | N/A | Pass | TRUE | ✓ | |
| 92 | 0040.0110 Secondary Audible Alarm Loud [19m 7s] | N/A | Pass | TRUE | ✓ | |
| 93 | 0040.0120 Secondary Audible Alarm Loud heard by HW (1) [1 | N/A | Pass | TRUE | ✓ | |
| 94 | 0040.0130 LCD Brightness [19m 13s] | N/A | Pass | TRUE | ✓ | |
| 95 | 0040.0140 LCD screen test [19m 14s] | N/A | Pass | TRUE | ✓ | |
| 96 | 0040.0150 120 VAC Power Source EPAP Current (1) [21m 46s] | A | 0,15 to 1,30 | 0,29 | ✓ | |
| 97 | 0040.0160 120 VAC Power Source IPAP Current (1) [21m 46s] | A | 0,25 to 1,60 | 0,40 | ✓ | |
| 98 | 0040.0170 Detach. Li-Ion Batt. Charge (at 75% Cap.) (1) | mA | GE 250,00 | 1907,00 | ✓ | |
| 99 | 0040.0180 Uninterrupted Operation (1) [21m 46s] | N/A | Pass | TRUE | ✓ | |
| 100 | 0040.0190 AC Power Source Recognition (1) [21m 46s] | N/A | EQ 120V AC Power | 120V AC Power | ✓ | |
| 101 | 0040.0200 AC Power Source LED ON (1) [21m 46s] | N/A | Pass | TRUE | ✓ | |
| 102 | 0040.0210 240 VAC Power Source EPAP Current (1) [22m 23s] | A | 0,07 to 0,80 | 0,13 | ✓ | |
| 103 | 0040.0220 240 VAC Power Source IPAP Current (1) [22m 23s] | A | 0,12 to 1,00 | 0,22 | ✓ | |
| 104 | 0040.0230 Uninterrupted Operation (1) [22m 23s] | N/A | Pass | TRUE | ✓ | |
| 105 | 0040.0240 AC Power Source Recognition (1) [22m 23s] | N/A | EQ 120V AC Power | 120V AC Power | ✓ | |

Test Started On 02/26/24 10:02:09Serial Number TV12003305BElapsed Test Time 50m 44sStatus **PASS**

Page 3 of 6

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Production mode.
 File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
 Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20