Test Software P/N 4102483 **Tested Part Number** 1054096

Test Software Name Trilogy FSA (PM TEST)

Software Revision 24.0.0.0 **Prepared By** P. Pascal

PHILIPS RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
71	0030.1030 Contr.Press.Ver.(37139): P1 at 45,67 (SP 46, 1,	cmH2O	44,67 to 46,67	45,99	✓	
72	0030.1040 Monit.Press.Ver.(35436): P2 at 45,67 (SP 46, 1,	cmH2O	44,67 to 46,67	45,84	✓	
73	0030.0164 Pprox Press. Ver.(36758): P2 at 45,67 (SP 46, 1	cmH2O	44,67 to 46,67	45,86	✓	
74	0030.1050 Contr.Press.Ver.(39530): P1 at 49,67 (SP 50, 1,	cmH2O	48,67 to 50,67	50,00	✓	
75	0030.1060 Monit.Press.Ver.(38242): P2 at 49,67 (SP 50, 1,	cmH2O	48,67 to 50,67	49,85	✓	
76	0030.0165 Pprox Press. Ver.(39191): P2 at 49,67 (SP 50, 1	cmH2O	48,67 to 50,67	49,78	✓	
77	0030.1070 Contr.Press.Ver.(48382): P1 at 64,34 (SP 65, 1,	cmH2O	63,34 to 65,34	64,65	✓	
78	0030.1080 Monit.Press.Ver.(48481): P2 at 64,34 (SP 65, 1,	cmH2O	63,34 to 65,34	64,56	✓	
79	0030.0166 Pprox Press. Ver.(48017): P2 at 64,34 (SP 65, 1	cmH2O	63,34 to 65,34	64,60	✓	
80	0030.1150 Atmospheric Pressure Verify: Patm at 28,05,28,0	inHg	27,45 to 28,65	27,93	✓	
81	0030.1250 Airstream Temperature (Tair) at 27,00 [28m 26s]	C deg	22,00 to 42,00	29,85	✓	

Test Started On 01/17/24 03:20:59

Serial Number TV110051733 **Elapsed Test Time** 28m 26s

Status **FAIL**

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Production mode.
File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20