Test Software P/N 4102483 **Tested Part Number** 1040000

Test Software Name Trilogy FSA (REPAIR TEST)

Software Revision 24.0.0.0 **Prepared By** P. Pascal

PHILIPS RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0020.0570 Pos Flow Cal: dP2 at 73,0 (Setpoint 75) [14m 28	ADC Cnts	LE 45981	45059	✓	
37	0020.0580 Pos Flow Cal: dP2 at 82,2 (Setpoint 85) [14m 28	ADC Cnts	LE 46952	45982	✓	
38	0020.0590 Pos Flow Cal: dP2 at 92,7 (Setpoint 95) [14m 28	ADC Cnts	LE 47938	46953	✓	
39	0020.0600 Pos Flow Cal: dP2 at 103,3 (Setpoint 105) [14m	ADC Cnts	LE 48754	47939	✓	
40	0020.0610 Pos Flow Cal: dP2 at 113,0 (Setpoint 115) [14m	ADC Cnts	LE 49434	48755	✓	
41	0020.0620 Pos Flow Cal: dP2 at 123,0 (Setpoint 125) [14m	ADC Cnts	LE 50146	49435	✓	
42	0020.0630 Pos Flow Cal: dP2 at 132,8 (Setpoint 135) [14m	ADC Cnts	LE 50750	50147	✓	
43	0020.0640 Pos Flow Cal: dP2 at 141,5 (Setpoint 145) [14m	ADC Cnts	LE 52698	50751	✓	
44	0020.0650 Pos Flow Cal: dP2 at 172,5 (Setpoint 170) [14m	ADC Cnts	LE 53587	52699	✓	
45	0020.0660 Pos Flow Cal: dP2 at 187,5 (Setpoint 190) [14m	ADC Cnts	LE 65435	53588	✓	
46	0020.0900 Atmospheric Press Sensors Cal Patm: Slope (m) [	inHg/ADC	EQ 0,000517	0,000517	✓	
47	0020.0910 Atmospheric Press Sensors Cal Patm: Intercept (	inHg	1,91 to 4,09	2,49	✓	
48	0020.0920 Write Sensors Cal Table [14m 52s]	N/A	Pass	TRUE	✓	
49	0020.0970 Write Pprox Cal Table [15m 8s]	N/A	Pass	TRUE	✓	
50	0020.0935 O2 Sensor Cal (not attempted for SW ver.14.2.05	N/A	None	None	✓	
51	0020.0980 dP Calibration: dP1 @ 10,00cmH2O (Setpoint 10,0	ADC Cnts	14026 to 67176	46855	✓	
52	0020.0981 dP Calibration: dP1 @ 4,00cmH2O (Setpoint 4,00)	ADC Cnts	195626 to 231476	212632	✓	
53	0020.0982 dP Calibration: dP1 @ 0,00cmH2O (Setpoint 0,00)	ADC Cnts	297270 to 358070	328856	✓	
54	0020.0983 dP Calibration: dP1 @ 4,00cmH2O (Setpoint 4,00)	ADC Cnts	426236 to 462086	446938	✓	
55	0020.0984 dP Calibration: dP1 @ 10,00cmH2O (Setpoint 10,0	ADC Cnts	590536 to 643686	619869	✓	
56	0020.1020 Write dP Cal Table [20m 33s]	N/A	Pass	TRUE	✓	
57	0020.1030 Write Image Table [20m 46s]	N/A	Pass	TRUE	✓	

Test Started On 03/05/24 11:50:19

Serial Number TV019111308 **Elapsed Test Time** 20m 46s

Status **PASS**

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Troubleshooting: Group Level mode.
File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 77D6, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20