Test Software P/N 4102483

**Tested Part Number** 1040000

Test Software Name Trilogy FSA (REPAIR TEST)



**Software Revision** 24.0.0.0

Prepared By P. Pascal

## Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0021.0570 O2 Positive Flow Cal: dP1 at 64,2 (SP 65 @81,1P.	ADC Cnts	LE 1875	1734	<b>√</b>	
37	0021.0580 O2 Positive Flow Cal: dP1 at 74,0 (SP 75 @80,4P	ADC Cnts	LE 2028	1876	<b>√</b>	
38	0021.0590 O2 Positive Flow Cal: dPl at 84,8 (SP 85 @79,6P	ADC Cnts	LE 2141	2030	<b>√</b>	
39	0021.0600 02 Positive Flow Cal: dP1 at 93,7 (SP 95 @78,9P	ADC Cnts	LE 2267	2142	<b>√</b>	
40	0021.0610 02 Positive Flow Cal: dP1 at 104,8 (SP 105 @77,	ADC Cnts	LE 2375	2268	<b>√</b>	
41	0021.0620 O2 Positive Flow Cal: dPl at 114,4 (SP 115 @77,	ADC Cnts	LE 2472	2376	<b>√</b>	
42	0021.0630 O2 Positive Flow Cal: dPl at 124,0 (SP 125 @76,	ADC Cnts	LE 2576	2473	<b>√</b>	
43	0021.0640 O2 Positive Flow Cal: dP1 at 134,0 (SP 135 @75,	ADC Cnts	LE 2676	2577	<b>√</b>	
44	0021.0650 O2 Positive Flow Cal: dP1 at 145,3 (SP 145 @74,	ADC Cnts	LE 2934	2677	<b>√</b>	
45	0021.0660 O2 Positive Flow Cal: dP1 at 174,2 (SP 175 @71,	ADC Cnts	LE 3063	2935	<b>√</b>	
46	0021.0670 O2 Positive Flow Cal: dP1 at 189,0 (SP 190 @68,	ADC Cnts	LE 3999	3064	<b>√</b>	
47	0021.0880 02 Sensor Cal (RANGE_FAILURE): 1 mV at ambient	N/A	Pass	FALSE		×
48	0021.0930 Write OBM Sensors Cal Table [16m 15s]	N/A	Pass	FALSE		×

**Test Started On** 06/19/24 02:53:57

Elapsed Test Time 16m 15s

Status FAIL

Serial Number TV01511120D

Page 2 **of** 2